

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

EPA Sample No.: SSTD020065 Date Analyzed: 01/14/2025

Lab File ID: BM049343.D Time Analyzed: 09:26

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	232920	7.522	944769	10.29	661003	14.16
UPPER LIMIT	465840	8.022	1889540	10.786	1322010	14.662
LOWER LIMIT	116460	7.022	472385	9.786	330502	13.662
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	174534	7.52	666169	10.29	446589	14.16

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.