

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.677	0.593	12.4	104	0.00
3 P	Chloromethane	0.733	0.629	14.2	102	0.00
4 C	Vinyl Chloride	0.737	0.629	14.7#	101	0.00
5 T	Bromomethane	0.445	0.364	18.2	96	0.00
6 T	Chloroethane	0.467	0.395	15.4	103	0.00
7 T	Trichlorofluoromethane	1.069	0.940	12.1	104	0.00
8 T	Diethyl Ether	0.369	0.316	14.4	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.538	10.6	109	0.00
10 T	Methyl Iodide	0.690	0.638	7.5	98	0.00
11 T	Tert butyl alcohol	0.092	0.076	17.4	90	0.00
12 CM	1,1-Dichloroethene	0.537	0.486	9.5#	101	0.00
13 T	Acrolein	0.126	0.124	1.6	97	0.00
14 T	Allyl chloride	0.871	0.775	11.0	101	0.00
15 T	Acrylonitrile	0.294	0.268	8.8	99	0.00
16 T	Acetone	0.261	0.223	14.6	98	0.00
17 T	Carbon Disulfide	1.652	1.366	17.3	102	0.00
18 T	Methyl Acetate	0.793	0.714	10.0	100	0.00
19 T	Methyl tert-butyl Ether	1.742	1.665	4.4	98	0.00
20 T	Methylene Chloride	0.646	0.555	14.1	97	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.512	10.6	102	0.00
22 T	Diisopropyl ether	1.933	1.817	6.0	98	0.00
23 T	Vinyl Acetate	1.353	1.339	1.0	99	0.00
24 P	1,1-Dichloroethane	1.178	1.040	11.7	98	0.00
25 T	2-Butanone	0.384	0.351	8.6	99	0.00
26 T	2,2-Dichloropropane	0.953	0.907	4.8	107	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.614	9.0	99	0.00
28 T	Bromochloromethane	0.548	0.484	11.7	98	0.00
29 T	Tetrahydrofuran	0.243	0.233	4.1	98	0.00
30 C	Chloroform	1.218	1.074	11.8#	101	0.00
31 T	Cyclohexane	1.024	0.878	14.3	110	0.00
32 T	1,1,1-Trichloroethane	1.068	0.947	11.3	103	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.804	0.4	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.347	0.352	-1.4	107	0.00
36 T	1,1-Dichloropropene	0.487	0.437	10.3	103	0.00
37 T	Ethyl Acetate	0.491	0.434	11.6	97	0.00
38 T	Carbon Tetrachloride	0.557	0.496	11.0	102	0.00
39 T	Methylcyclohexane	0.457	0.455	0.4	107	0.00
40 TM	Benzene	1.463	1.336	8.7	100	0.00
41 T	Methacrylonitrile	0.256	0.241	5.9	97	0.00
42 TM	1,2-Dichloroethane	0.551	0.501	9.1	100	0.00
43 T	Isopropyl Acetate	0.787	0.716	9.0	96	0.00
44 TM	Trichloroethene	0.341	0.297	12.9	100	0.00
45 C	1,2-Dichloropropane	0.374	0.334	10.7#	98	0.00
46 T	Dibromomethane	0.270	0.238	11.9	97	0.00
47 T	Bromodichloromethane	0.549	0.502	8.6	98	0.00
48 T	Methyl methacrylate	0.354	0.354	0.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085445.D
 Acq On : 14 Jan 2025 18:06
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN011425

Quant Time: Jan 15 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	81	0.00
50 S	Toluene-d8	1.232	1.289	-4.6	111	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.449	1.8	99	0.00
52 CM	Toluene	0.848	0.815	3.9#	102	0.00
53 T	t-1,3-Dichloropropene	0.519	0.496	4.4	98	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.530	4.3	98	0.00
55 T	1,1,2-Trichloroethane	0.335	0.310	7.5	99	0.00
56 T	Ethyl methacrylate	0.469	0.492	-4.9	99	0.00
57 T	1,3-Dichloropropane	0.583	0.547	6.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.218	-2.3	99	0.00
59 T	2-Hexanone	0.321	0.323	-0.6	99	0.00
60 T	Dibromochloromethane	0.405	0.372	8.1	98	0.00
61 T	1,2-Dibromoethane	0.334	0.299	10.5	98	0.00
62 S	4-Bromofluorobenzene	0.422	0.446	-5.7	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.341	0.306	10.3	102	0.00
65 PM	Chlorobenzene	1.095	1.006	8.1	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.359	10.7	99	0.00
67 C	Ethyl Benzene	1.784	1.798	-0.8#	104	0.00
68 T	m/p-Xylenes	0.659	0.681	-3.3	104	0.00
69 T	o-Xylene	0.630	0.656	-4.1	104	0.00
70 T	Styrene	1.043	1.108	-6.2	102	0.00
71 P	Bromoform	0.288	0.283	1.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.375	3.371	0.1	106	0.00
74 T	N-amyl acetate	1.517	1.478	2.6	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.040	12.8	98	0.00
76 T	1,2,3-Trichloropropane	1.016	0.851	16.2	103	0.00
77 T	Bromobenzene	0.882	0.814	7.7	103	0.00
78 T	n-propylbenzene	3.995	4.075	-2.0	106	0.00
79 T	2-Chlorotoluene	2.586	2.475	4.3	104	0.00
80 T	1,3,5-Trimethylbenzene	2.787	2.877	-3.2	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.370	1.6	103	0.00
82 T	4-Chlorotoluene	2.574	2.511	2.4	105	0.00
83 T	tert-Butylbenzene	2.341	2.387	-2.0	106	0.00
84 T	1,2,4-Trimethylbenzene	2.778	2.885	-3.9	103	0.00
85 T	sec-Butylbenzene	3.244	3.360	-3.6	106	0.00
86 T	p-Isopropyltoluene	2.631	2.777	-5.5	105	0.00
87 T	1,3-Dichlorobenzene	1.624	1.491	8.2	103	0.00
88 T	1,4-Dichlorobenzene	1.683	1.484	11.8	103	0.00
89 T	n-Butylbenzene	2.327	2.354	-1.2	103	0.00
90 T	Hexachloroethane	0.618	0.559	9.5	105	0.00
91 T	1,2-Dichlorobenzene	1.620	1.461	9.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.201	7.8	98	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.689	8.5	95	0.00
94 T	Hexachlorobutadiene	0.400	0.337	15.8	99	0.00
95 T	Naphthalene	2.246	2.117	5.7	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.679	10.9	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6