

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085440.D
 Acq On : 14 Jan 2025 15:43
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	200796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	340593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143526	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	60593	18.694	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	37.380%#
35) Dibromofluoromethane	8.165	113	45655	19.322	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	38.640%#
50) Toluene-d8	10.559	98	164391	19.581	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	39.160%#
62) 4-Bromofluorobenzene	12.847	95	55798	19.430	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	38.860%#

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	54666	20.107	ug/l	98
3) Chloromethane	2.359	50	58426	19.848	ug/l	100
4) Vinyl Chloride	2.512	62	58403	19.739	ug/l	99
5) Bromomethane	2.954	94	36440	20.389	ug/l	94
6) Chloroethane	3.118	64	37594	20.041	ug/l	93
7) Trichlorofluoromethane	3.501	101	88121	20.526	ug/l	99
8) Diethyl Ether	3.965	74	29138	19.647	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	48933	20.237	ug/l	99
10) Methyl Iodide	4.589	142	56567	20.428	ug/l	93
11) Tert butyl alcohol	5.506	59	37984	102.342	ug/l	100
12) 1,1-Dichloroethene	4.342	96	44670	20.728	ug/l	95
13) Acrolein	4.171	56	49649	97.995	ug/l	99
14) Allyl chloride	5.024	41	67705	19.362	ug/l	96
15) Acrylonitrile	5.712	53	121626	103.171	ug/l	99
16) Acetone	4.424	43	101065	96.502	ug/l	96
17) Carbon Disulfide	4.712	76	132300	19.939	ug/l	98
18) Methyl Acetate	5.018	43	60873	19.115	ug/l	99
19) Methyl tert-butyl Ether	5.795	73	148847	21.272	ug/l	100
20) Methylene Chloride	5.271	84	52813	20.370	ug/l	97
21) trans-1,2-Dichloroethene	5.789	96	46115	20.023	ug/l	98
22) Diisopropyl ether	6.671	45	166719	21.480	ug/l	97
23) Vinyl Acetate	6.594	43	588567m	107.984	ug/l	
24) 1,1-Dichloroethane	6.565	63	96875	20.469	ug/l #	96
25) 2-Butanone	7.477	43	159952	103.786	ug/l	98
26) 2,2-Dichloropropane	7.483	77	79145	20.684	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	57454	21.180	ug/l	98
28) Bromochloromethane	7.812	49	41181	18.703	ug/l	99
29) Tetrahydrofuran	7.836	42	104792	107.247	ug/l	99
30) Chloroform	7.959	83	99655	20.374	ug/l	99
31) Cyclohexane	8.253	56	82980	20.173	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	87659	20.431	ug/l	98
36) 1,1-Dichloropropene	8.371	75	67949	20.490	ug/l	98
37) Ethyl Acetate	7.553	43	66443	19.850	ug/l	96
38) Carbon Tetrachloride	8.359	117	78842	20.767	ug/l	96
39) Methylcyclohexane	9.594	83	65012	20.862	ug/l	96
40) Benzene	8.600	78	207985	20.873	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085440.D
 Acq On : 14 Jan 2025 15:43
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Jan 15 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	36578	21.004	ug/l	99
42) 1,2-Dichloroethane	8.665	62	78402	20.891	ug/l	99
43) Isopropyl Acetate	8.683	43	109131	20.347	ug/l	98
44) Trichloroethene	9.347	130	48009	20.699	ug/l	96
45) 1,2-Dichloropropane	9.618	63	52833	20.758	ug/l	93
46) Dibromomethane	9.700	93	37721	20.540	ug/l	98
47) Bromodichloromethane	9.882	83	78918	21.093	ug/l	96
48) Methyl methacrylate	9.677	41	50224	20.808	ug/l	96
49) 1,4-Dioxane	9.688	88	17140	421.686	ug/l	95
51) 4-Methyl-2-Pentanone	10.441	43	336911	108.249	ug/l	100
52) Toluene	10.630	92	125172	21.680	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	74129	20.965	ug/l	96
54) cis-1,3-Dichloropropene	10.306	75	81897	21.684	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	48039	21.025	ug/l	97
56) Ethyl methacrylate	10.871	69	70934	19.649	ug/l	99
57) 1,3-Dichloropropane	11.159	76	84210	21.196	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	153681	105.991	ug/l	100
59) 2-Hexanone	11.194	43	240595	109.863	ug/l	98
60) Dibromochloromethane	11.353	129	56069	20.326	ug/l	96
61) 1,2-Dibromoethane	11.465	107	48465	21.309	ug/l	96
64) Tetrachloroethene	11.100	164	43243	21.408	ug/l	96
65) Chlorobenzene	11.888	112	136809	21.077	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	48370	20.304	ug/l	98
67) Ethyl Benzene	11.959	91	229978	21.753	ug/l	98
68) m/p-Xylenes	12.071	106	177755	45.494	ug/l	100
69) o-Xylene	12.394	106	84548	22.643	ug/l	98
70) Styrene	12.412	104	140581	22.749	ug/l	99
71) Bromoform	12.576	173	36961	21.683	ug/l #	96
73) Isopropylbenzene	12.694	105	211354	21.819	ug/l	100
74) N-ethyl acetate	12.494	43	94292	21.656	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	68151	19.917	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	64180m	22.036	ug/l	
77) Bromobenzene	12.976	156	52022	20.556	ug/l	99
78) n-propylbenzene	13.035	91	252610	22.028	ug/l	100
79) 2-Chlorotoluene	13.123	91	158511	21.358	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	180294	22.533	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	22178m	20.569	ug/l	
82) 4-Chlorotoluene	13.218	91	161880	21.906	ug/l	100
83) tert-Butylbenzene	13.435	119	147541	21.958	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	183144	22.965	ug/l	99
85) sec-Butylbenzene	13.612	105	209133	22.457	ug/l	100
86) p-Isopropyltoluene	13.729	119	168888	21.661	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	97652	20.952	ug/l	100
88) 1,4-Dichlorobenzene	13.806	146	98321	20.351	ug/l	99
89) n-Butylbenzene	14.053	91	142673	21.355	ug/l	99
90) Hexachloroethane	14.329	117	35205	19.858	ug/l	98
91) 1,2-Dichlorobenzene	14.100	146	94954	20.425	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12859	20.580	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	45878	21.233	ug/l	99
94) Hexachlorobutadiene	15.500	225	22719	19.805	ug/l	95
95) Naphthalene	15.635	128	134793	20.906	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	45468	20.797	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 168 Resp: 200796
Ion Ratio Lower Upper
168 100
99 62.2 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 20.107 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 85 Resp: 54666
Ion Ratio Lower Upper
85 100
87 32.5 16.7 50.1









#3
Chloromethane
Concen: 19.848 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 58426
Ion Ratio Lower Upper
50 100
52 34.0 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#4
Vinyl Chloride
Concen: 19.739 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 62 Resp: 58403
Ion Ratio Lower Upper
62 100
64 30.7 24.8 37.2









#5
Bromomethane
Concen: 20.389 ug/l
RT: 2.954 min Scan# 170
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 94 Resp: 36440
Ion Ratio Lower Upper
94 100
96 94.9 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#6
Chloroethane
Concen: 20.041 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 64 Resp: 37594
Ion Ratio Lower Upper
64 100
66 34.3 24.6 36.8






#7
 Trichlorofluoromethane
 Concen: 20.526 ug/l
 RT: 3.501 min Scan# 263
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020






Tgt Ion:101 Resp: 88121
 Ion Ratio Lower Upper
 101 100
 103 65.3 51.4 77.2

Manual Integrations




APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#8
 Diethyl Ether
 Concen: 19.647 ug/l
 RT: 3.965 min Scan# 342
 Delta R.T. 0.012 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 74 Resp: 29138
 Ion Ratio Lower Upper
 74 100
 45 101.8 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.237 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 48933
Ion Ratio Lower Upper
101 100
85 46.6 37.8 56.8
151 72.1 58.8 88.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 20.428 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:142 Resp: 56567
Ion Ratio Lower Upper
142 100
127 44.6 40.3 60.5
141 14.8 12.8 19.2





#11

Tert butyl alcohol

Concen: 102.342 ug/l

RT: 5.506 min Scan# 605

Delta R.T. -0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 59 Resp: 37984

Ion Ratio Lower Upper

59 100





57 10.7 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 20.728 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 96 Resp: 44670


Ion Ratio Lower Upper

96 100

61 170.3 141.0 211.4

98 56.8 49.0 73.6





#13

Acrolein

Concen: 97.995 ug/l

RT: 4.171 min Scan# 37

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 49649

Ion Ratio Lower Upper

56 100





55 71.4 56.3 84.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#14

Allyl chloride

Concen: 19.362 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 41 Resp: 67705


Ion Ratio Lower Upper

41 100

39 85.0 64.4 96.6


76 38.6 30.5 45.7





#15
Acrylonitrile
Concen: 103.171 ug/l
RT: 5.712 min Scan# 639
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion: 53 Resp: 121626
Ion Ratio Lower Upper
53 100
52 81.1 65.5 98.3
51 36.1 29.8 44.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 96.502 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43



Tgt Ion: 43 Resp: 101065
Ion Ratio Lower Upper
43 100
58 27.7 23.8 35.6









#17
Carbon Disulfide
Concen: 19.939 ug/l
RT: 4.712 min Scan# 468
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 76 Resp: 132300
Ion Ratio Lower Upper
76 100
78 9.3 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#18
Methyl Acetate
Concen: 19.115 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 60873
Ion Ratio Lower Upper
43 100
74 24.0 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 21.272 ug/l
RT: 5.795 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 73 Resp: 148847
Ion Ratio Lower Upper
73 100
57 22.1 17.8 26.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#20
Methylene Chloride
Concen: 20.370 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 84 Resp: 52813
Ion Ratio Lower Upper
84 100
49 129.2 104.1 156.1
51 42.2 31.0 46.4
86 61.6 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 20.023 ug/l

RT: 5.789 min Scan# 651

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 96 Resp: 46115

Ion Ratio Lower Upper

96 100

61 146.1 115.7 173.5





98 64.5 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether

Concen: 21.480 ug/l

RT: 6.671 min Scan# 802

Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 45 Resp: 166719




Ion Ratio Lower Upper


45 100

43 54.3 45.0 67.6

87 25.3 21.8 32.8

59 10.8 10.2 15.2





#23

Vinyl Acetate

Concen: 107.984 ug/l m

RT: 6.594 min Scan# 789

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 588567

Ion Ratio Lower Upper

43 100





86 10.1 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 20.469 ug/l

RT: 6.565 min Scan# 784

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43




Tgt Ion: 63 Resp: 96875


Ion Ratio Lower Upper

63 100

98 7.4 3.3 9.8

100 6.0 2.0 6.0#









#25
2-Butanone
Concen: 103.786 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 159952
Ion Ratio Lower Upper
43 100
72 24.1 20.2 30.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#26
2,2-Dichloropropane
Concen: 20.684 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 77 Resp: 79145
Ion Ratio Lower Upper
77 100
97 22.0 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 21.180 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Tgt Ion: 96 Resp: 57454
 Ion Ratio Lower Upper
 96 100
 61 153.0 0.0 311.8
 98 63.7 0.0 126.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#28
 Bromochloromethane
 Concen: 18.703 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Tgt Ion: 49 Resp: 41181
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 69.3 55.0 82.4









#29
Tetrahydrofuran
Concen: 107.247 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 42 Resp: 104792
Ion Ratio Lower Upper
42 100
72 42.6 34.2 51.2
71 41.4 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#30
Chloroform
Concen: 20.374 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 83 Resp: 99655
Ion Ratio Lower Upper
83 100
85 65.8 51.8 77.6









#31
Cyclohexane
Concen: 20.173 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:	56	Resp:	82980
Ion	Ratio	Lower	Upper
56	100		
69	31.2	22.2	33.4
84	80.8	66.4	99.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#32
1,1,1-Trichloroethane
Concen: 20.431 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:	97	Resp:	87659
Ion	Ratio	Lower	Upper
97	100		
99	64.2	49.8	74.6
61	50.3	41.4	62.2









#33
1,2-Dichloroethane-d4
Concen: 18.694 ug/l
RT: 8.571 min Scan# 1125
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 65 Resp: 60593
Ion Ratio Lower Upper
65 100
67 51.2 0.0 101.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion:114 Resp: 340593
Ion Ratio Lower Upper
114 100
63 21.6 0.0 47.6
88 16.0 0.0 32.6





#35
Dibromofluoromethane
Concen: 19.322 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 113 Resp: 45655
Ion Ratio Lower Upper
113 100
111 100.6 82.7 124.1
192 17.9 14.3 21.5

Manual Integrations
APPROVED



Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 20.490 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43



Tgt Ion: 75 Resp: 67949
Ion Ratio Lower Upper
75 100
110 31.8 16.5 49.5
77 31.1 24.4 36.6









#37
Ethyl Acetate
Concen: 19.850 ug/l
RT: 7.553 min Scan# 952
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43	Resp: 66443
Ion Ratio	Lower Upper
43 100	
61 11.9	10.9 16.3
70 10.6	7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#38
Carbon Tetrachloride
Concen: 20.767 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 117	Resp: 78842
Ion Ratio	Lower Upper
117 100	
119 98.9	75.4 113.2
121 28.9	24.6 37.0









#39
Methylcyclohexane
Concen: 20.862 ug/l
RT: 9.594 min Scan# 1299
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 83 Resp: 65012
Ion Ratio Lower Upper
83 100
55 83.4 62.6 94.0
98 46.4 37.7 56.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#40
Benzene
Concen: 20.873 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 78 Resp: 207985
Ion Ratio Lower Upper
78 100
77 24.4 19.0 28.6









#41
Methacrylonitrile
Concen: 21.004 ug/l
RT: 7.771 min Scan# 989
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 41	Resp: 36578
Ion Ratio	Lower Upper
41 100	
39 56.6	46.0 69.0
67 70.3	57.4 86.2
52 32.4	25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#42
1,2-Dichloroethane
Concen: 20.891 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 62	Resp: 78402
Ion Ratio	Lower Upper
62 100	
98 8.3	0.0 17.0





#43

Isopropyl Acetate

Concen: 20.347 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 43	Resp: 109131
Ion Ratio	Lower Upper
43	100
61	27.2 20.7 31.1
87	12.3 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 20.699 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion: 130	Resp: 48009
Ion Ratio	Lower Upper
130	100
95	101.5 0.0 195.8









#45
1,2-Dichloropropane
Concen: 20.758 ug/l
RT: 9.618 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 63 Resp: 52833
Ion Ratio Lower Upper
63 100
65 28.3 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#46
Dibromomethane
Concen: 20.540 ug/l
RT: 9.700 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 93 Resp: 37721
Ion Ratio Lower Upper
93 100
95 82.0 64.7 97.1
174 88.4 69.0 103.4





#47
 Bromodichloromethane
 Concen: 21.093 ug/l
 RT: 9.882 min Scan# 1348
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC020

Tgt Ion:	83	Resp:	78918
Ion	Ratio	Lower	Upper
83	100		
85	67.8	51.2	76.8
127	8.6	6.5	9.7




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#48
 Methyl methacrylate
 Concen: 20.808 ug/l
 RT: 9.677 min Scan# 1313
 Delta R.T. -0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43


Tgt Ion:	41	Resp:	50224
Ion	Ratio	Lower	Upper
41	100		
69	84.3	64.7	97.1
39	64.2	49.0	73.6





#49
1,4-Dioxane
Concen: 421.686 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 88 Resp: 17140
Ion Ratio Lower Upper
88 100
43 34.9 26.6 39.8
58 69.0 59.5 89.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#50
Toluene-d8
Concen: 19.581 ug/l
RT: 10.559 min Scan# 1463
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43



Tgt Ion: 98 Resp: 164391
Ion Ratio Lower Upper
98 100
100 64.4 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 108.249 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 336911
Ion Ratio Lower Upper
43 100
58 38.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 21.680 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion: 92 Resp: 125172
Ion Ratio Lower Upper
92 100
91 175.1 139.2 208.8





#53
 t-1,3-Dichloropropene
 Concen: 20.965 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020




Tgt Ion: 75 Resp: 74129
 Ion Ratio Lower Upper
 75 100
 77 31.4 23.5 35.3

Manual Integrations
 APPROVED



Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025




#54
 cis-1,3-Dichloropropene
 Concen: 21.684 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.000 min
 Lab File: VN085440.D
 Acq: 14 Jan 2025 15:43




Tgt Ion: 75 Resp: 81897
 Ion Ratio Lower Upper
 75 100
 77 32.4 25.0 37.4
 39 54.4 43.1 64.7





#55
1,1,2-Trichloroethane
Concen: 21.025 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion	Ratio	Lower	Upper
97	100		
83	84.1	69.0	103.6
85	54.6	44.6	66.8
99	64.8	48.6	72.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 19.649 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Tgt Ion	Ratio	Lower	Upper
69	100		
41	67.6	54.6	82.0
39	41.7	32.4	48.6





#57
1,3-Dichloropropane
Concen: 21.196 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 76 Resp: 84210
Ion Ratio Lower Upper
76 100
78 32.7 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#58
2-Chloroethyl Vinyl ether
Concen: 105.991 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 63 Resp: 153681
Ion Ratio Lower Upper
63 100
106 27.2 21.6 32.4









#59
2-Hexanone
Concen: 109.863 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 43 Resp: 240595
Ion Ratio Lower Upper
43 100
58 51.1 26.2 78.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#60
Dibromochloromethane
Concen: 20.326 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:129 Resp: 56069
Ion Ratio Lower Upper
129 100
127 80.4 38.6 115.8









#61
1,2-Dibromoethane
Concen: 21.309 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 107 Resp: 48465
Ion Ratio Lower Upper
107 100
109 90.6 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 19.430 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 95 Resp: 55798
Ion Ratio Lower Upper
95 100
174 75.2 0.0 145.0
176 73.2 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:	117	Resp:	296298
Ion	Ratio	Lower	Upper
117	100		
82	59.3	48.6	72.8
119	32.2	26.6	39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#64
Tetrachloroethene
Concen: 21.408 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:	164	Resp:	43243
Ion	Ratio	Lower	Upper
164	100		
166	130.6	103.4	155.2
129	107.4	79.2	118.8
131	94.7	77.1	115.7









#65
Chlorobenzene
Concen: 21.077 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:112 Resp: 136809
Ion Ratio Lower Upper
112 100
114 32.9 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.304 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:131 Resp: 48370
Ion Ratio Lower Upper
131 100
133 96.9 47.4 142.3
119 67.2 33.1 99.5









#67
Ethyl Benzene
Concen: 21.753 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 229978
Ion Ratio Lower Upper
91 100
106 31.1 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#68
m/p-Xylenes
Concen: 45.494 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:106 Resp: 177755
Ion Ratio Lower Upper
106 100
91 209.3 167.7 251.5









#69
o-Xylene
Concen: 22.643 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:106 Resp: 84548
Ion Ratio Lower Upper
106 100
91 217.0 110.4 331.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#70
Styrene
Concen: 22.749 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:104 Resp: 140581
Ion Ratio Lower Upper
104 100
78 54.4 42.5 63.7
103 54.8 43.8 65.8





#71

Bromoform

Concen: 21.683 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	173	Resp:	36961
Ion Ratio	Lower	Upper	
173	100		
175	51.7	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	152	Resp:	143526
Ion Ratio	Lower	Upper	
152	100		
115	61.4	31.1	93.3
150	163.9	0.0	343.6









#73
Isopropylbenzene
Concen: 21.819 ug/l
RT: 12.694 min Scan# 1825
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 105 Resp: 211354
Ion Ratio Lower Upper
105 100
120 25.5 12.8 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#74
N-ethyl acetate
Concen: 21.656 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 94292
Ion Ratio Lower Upper
43 100
70 41.4 34.0 51.0
55 27.9 21.4 32.2
61 23.1 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 19.917 ug/l

RT: 12.935 min Scan# 18

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	83	Resp:	68151
Ion Ratio	Lower	Upper	
83	100		
131	9.9	4.8	14.4
85	65.5	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 22.036 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	75	Resp:	64180
Ion Ratio	Lower	Upper	
75	100		
77	186.8	109.7	329.2





#77

Bromobenzene

Concen: 20.556 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	156	Resp:	52022
Ion Ratio	Lower	Upper	
156	100		
77	230.4	114.1	342.4
158	97.0	48.9	146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene

Concen: 22.028 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. 0.006 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	91	Resp:	252610
Ion Ratio	Lower	Upper	
91	100		
120	21.8	10.9	32.6









#79
2-Chlorotoluene
Concen: 21.358 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 91 Resp: 158511
Ion Ratio Lower Upper
91 100
126 31.5 15.7 47.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#80
1,3,5-Trimethylbenzene
Concen: 22.533 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion:105 Resp: 180294
Ion Ratio Lower Upper
105 100
120 47.9 23.9 71.7





#81
trans-1,4-Dichloro-2-butene
Concen: 20.569 ug/l m
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion: 75 Resp: 22178
Ion Ratio Lower Upper
75 100
53 107.2 95.6 143.4
89 46.9 37.0 55.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 21.906 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43



Tgt Ion: 91 Resp: 161880
Ion Ratio Lower Upper
91 100
126 31.6 15.9 47.7









#83
tert-Butylbenzene
Concen: 21.958 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:119 Resp: 147541
Ion Ratio Lower Upper
119 100
91 69.5 35.1 105.3
134 25.5 12.0 36.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#84
1,2,4-Trimethylbenzene
Concen: 22.965 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:105 Resp: 183144
Ion Ratio Lower Upper
105 100
120 42.8 21.6 65.0









#85
sec-Butylbenzene
Concen: 22.457 ug/l
RT: 13.612 min Scan# 1982
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:105 Resp: 209133
Ion Ratio Lower Upper
105 100
134 19.2 9.7 28.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#86
p-Isopropyltoluene
Concen: 21.661 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion:119 Resp: 168888
Ion Ratio Lower Upper
119 100
134 26.4 12.7 38.0
91 25.5 12.7 38.1





#87
1,3-Dichlorobenzene
Concen: 20.952 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:146 Resp: 97652
Ion Ratio Lower Upper
146 100
111 42.9 21.4 64.3
148 64.3 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 20.351 ug/l
RT: 13.806 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43



Tgt Ion:146 Resp: 98321
Ion Ratio Lower Upper
146 100
111 42.2 21.3 63.7
148 65.1 32.4 97.0









#89
n-Butylbenzene
Concen: 21.355 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 142673
Ion Ratio Lower Upper
91 100
92 51.0 25.8 77.3
134 23.0 11.7 35.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90
Hexachloroethane
Concen: 19.858 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion:117 Resp: 35205
Ion Ratio Lower Upper
117 100
201 68.8 33.7 101.0





#91
1,2-Dichlorobenzene
Concen: 20.425 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion:146 Resp: 94954
Ion Ratio Lower Upper
146 100
111 43.5 21.7 65.1
148 63.5 31.4 94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.580 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 12859
Ion Ratio Lower Upper
75 100
155 72.7 36.4 109.2
157 90.9 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 21.233 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	180	Resp:	45878
Ion Ratio	Lower	Upper	
180	100		
182	94.2	46.8	140.3
145	31.4	16.0	48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 19.805 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	225	Resp:	22719
Ion Ratio	Lower	Upper	
225	100		
223	64.3	30.7	92.1
227	66.3	30.9	92.5





#95

Naphthalene

Concen: 20.906 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	128	Resp:	134793
Ion Ratio	Lower	Upper	
128	100		
127	13.3	10.6	16.0
129	10.6	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 20.797 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085440.D

Acq: 14 Jan 2025 15:43

Tgt Ion:	180	Resp:	45468
Ion Ratio	Lower	Upper	
180	100		
182	97.9	47.4	142.2
145	35.7	16.9	50.7

