

Initial and Continuing Calibration Blank Summary

Client: Dewberry Engineers Inc.

SDG No.: Q1061

Project: PennDOT - Onvo Travel Plaza #50040379

RunNo.: LB134248

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/10/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/10/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/10/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/10/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/10/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/10/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/10/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/10/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/10/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/10/2025
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/10/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/10/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/10/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/10/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/10/2025
Sample ID: CCB4							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/10/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/10/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/10/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/10/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/10/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/10/2025
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	12/18/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	12/18/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	12/18/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	12/18/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	12/18/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	12/18/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	12/18/2024

Initial and Continuing Calibration Blank Summary

Client: Dewberry Engineers Inc.

SDG No.: Q1061

Project: PennDOT - Onvo Travel Plaza #50040379

RunNo.: LB134293

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.011	0.0250	J	0.0047	0.05	01/15/2025
Sample ID: CCB1 Phosphorus, Total	mg/L	0.011	0.0250	J	0.0047	0.05	01/15/2025
Sample ID: CCB2 Phosphorus, Total	mg/L	0.011	0.0250	J	0.0047	0.05	01/15/2025
Sample ID: CCB3 Phosphorus, Total	mg/L	0.011	0.0250	J	0.0047	0.05	01/15/2025