

Duplicate Sample Summary

Client:	Dewberry Engineers Inc.	SDG No.:	Q1061
Project:	PennDOT - Onvo Travel Plaza #50040379	Sample ID:	Q1061-01
Client ID:	MW-1DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.26		0.26		1	0.39		01/15/2025

Duplicate Sample Summary

Client: Dewberry Engineers Inc.	SDG No.: Q1061
Project: PennDOT - Onvo Travel Plaza #50040379	Sample ID: Q1061-01
Client ID: MW-1MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.20		2.20		1	0		01/10/2025
Nitrate	mg/L	+/-20	2.50		2.50		1	0		01/10/2025
Nitrite	mg/L	+/-20	3.00		3.00		1	0		01/10/2025
Chloride	mg/L	+/-20	32.1	OR	32.1	OR	1	0		01/10/2025
Bromide	mg/L	+/-20	10.1		10.0		1	1		01/10/2025
Sulfate	mg/L	+/-20	24.2		24.0		1	1		01/10/2025
Orthophosphate as P	mg/L	+/-20	4.90		4.70		1	4		01/10/2025
Phosphorus, Total	mg/L	+/-20	0.73		0.72		1	0.41		01/15/2025

Duplicate Sample Summary

Client: Dewberry Engineers Inc. Project: PennDOT - Onvo Travel Plaza #50040379 Client ID: 1 EFFLUENT COMPOSITEDUP	SDG No.: Q1061 Sample ID: Q1070-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	4120		4210		1	2.19		01/10/2025