

Laboratory Control Sample Summary

Client: Dewberry Engineers Inc.

SDG No.: Q1061

Project: PennDOT - Onvo Travel Plaza #50040379

Run No.: LB134233

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134233BS							
TDS	mg/L	100	95.0		95	1	90-110	01/10/2025

Laboratory Control Sample Summary

Client: Dewberry Engineers Inc.

SDG No.: Q1061

Project: PennDOT - Onvo Travel Plaza #50040379

Run No.: LB134248

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134248BSW							
Bromide	mg/L	10	10.2		102	1	90-110	01/10/2025
Chloride	mg/L	3	3.00		100	1	90-110	01/10/2025
Fluoride	mg/L	2	2.00		100	1	90-110	01/10/2025
Nitrite	mg/L	3	3.00		100	1	90-110	01/10/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	01/10/2025
Sulfate	mg/L	15	15.5		103	1	90-110	01/10/2025
Orthophosphate as P	mg/L	5	5.30		106	1	90-110	01/10/2025

Laboratory Control Sample Summary

Client: Dewberry Engineers Inc.

SDG No.: Q1061

Project: PennDOT - Onvo Travel Plaza #50040379

Run No.: LB134293

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166041BS							
Phosphorus, Total	mg/L	0.50	0.48		96	1	90-110	01/15/2025