

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Dewberry Engineers Inc.

**SDG No.:** Q1061

**Contract:** DEWB01

**Lab Code:** CHEM

**Case No.:** Q1061

**SAS No.:** Q1061

| Analyte    | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|------------|-------|------------|--------|---|---------------|----------------------|---|
| PB166036BS |       |            |        |   |               |                      |   |
| Potassium  | ug/L  | 5000       | 4580   |   | 92            | 85 - 115             | P |
| Sodium     | ug/L  | 1500       | 1420   |   | 95            | 85 - 115             | P |