

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1062 SAS No.: Q1062 SDG No.: Q1062

Instrument ID: BNA_N Calibration Date/Time: 01/10/2025 21:20

Lab File ID: BN035912.D Init. Calib. Date(s): 01/02/2025 01/02/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:28 15:04

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.536	0.515		-3.9	50.0
Fluoranthene-d10	0.993	0.961		-3.2	50.0
2-Fluorophenol	0.981	0.915		-6.7	50.0
Phenol-d6	1.219	1.137		-6.8	50.0
Nitrobenzene-d5	0.317	0.351		10.7	50.0
2-Fluorobiphenyl	1.755	1.801		2.6	50.0
2,4,6-Tribromophenol	0.192	0.220		14.6	50.0
Terphenyl-d14	0.797	0.765		-4.0	50.0
1,4-Dioxane	0.397	0.410		3.3	50.0

All other compounds must meet a minimum RRF of 0.010.