

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1062 SDG NO.: Q1062

Lab File ID: BN035913.D

DFTPP Injection Date: 01/13/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:28

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	52
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	46.2
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	47.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.1
275	10.0 - 60.0% of mass 198	24.6
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	9.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.2 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035914.D	01/13/2025	10:07
PB166005BL	PB166005BL	BN035915.D	01/13/2025	10:44
PB166005BS	PB166005BS	BN035917.D	01/13/2025	11:56
SSTDCCC0.4EC	SSTDCCC0.4	BN035918.D	01/13/2025	12:37