

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1062 SAS No.: Q1062 SDG NO.: Q1062
Lab File ID: VN085447.D Date Analyzed: 01/15/2025
Instrument ID: MSVOA_N Time Analyzed: 09:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	218273	8.22	358609	9.09	316952	11.86
UPPER LIMIT	436546	8.718	717218	9.594	633904	12.359
LOWER LIMIT	109137	7.718	179305	8.594	158476	11.359
EPA SAMPLE NO.						
BP-VPB-190A-TB-20250107	196038	8.22	355455	9.09	304386	11.87
VPB190A-HYD-20250107	183777	8.22	332391	9.10	293220	11.87
BP-VPB-190A-EB-20250107	185335	8.22	337105	9.10	298295	11.87
BP-VPB-190A-GW-898-900	175474	8.22	316452	9.10	274227	11.87
VN0115WBL01	195111	8.22	360723	9.10	319974	11.86
VN0115WBS01	208834	8.22	350038	9.09	301633	11.87
VN0115WBSD01	188898	8.22	323122	9.10	286298	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.