

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1062 SAS No.: Q1062 SDG NO.: Q1062
 Lab File ID: VN085447.D Date Analyzed: 01/15/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	158347	13.788				
UPPER LIMIT	316694	14.288				
LOWER LIMIT	79173.5	13.288				
EPA SAMPLE NO.						
BP-VPB-190A-TB-20250107	119672	13.79				
VPB190A-HYD-20250107	116419	13.79				
BP-VPB-190A-EB-20250107	117680	13.79				
BP-VPB-190A-GW-898-900	107831	13.79				
VN0115WBL01	130534	13.79				
VN0115WBS01	141189	13.79				
VN0115WBSD01	136502	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.