

Initial and Continuing Calibration Verification

Client: Ardmore Chemical

SDG No.: Q1066

Project: PVSC Monthly 2025

RunNo.: LB134287

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.097	0.099	98	85-115	01/14/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	01/14/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	01/14/2025

Initial and Continuing Calibration Verification

Client: Ardmore Chemical

SDG No.: Q1066

Project: PVSC Monthly 2025

RunNo.: LB134287

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------