

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q1066

Project: PVSC Monthly 2025

Run No.: LB134238

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134238BS							
BOD5	mg/L	198	201		102	1	84.6-115.4	01/10/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q1066

Project: PVSC Monthly 2025

Run No.: LB134244

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134244BS							
TSS	mg/L	550	520		94	1	90-110	01/13/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q1066

Project: PVSC Monthly 2025

Run No.: LB134287

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166058BS							
Cyanide	mg/L	0.1	0.098		98	1	85-115	01/14/2025