

LB134431


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

38677



A= 0.0000e+000

B= 2.5935e-004

C= -3.2280e-002

Rho= 0.9998673

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/0/0
0.0	0.000	-0.014	-0.014	69	0.000	69					1
0.05	0.050	0.036	-0.014	265	0.0 %	265	0				-28
0.2	0.200	0.157	-0.043	728	0.0 %	728					-22
2.5	2.500	2.555	0.055	9976	0.0 %	9976					2
5.0	5.000	5.109	0.109	19825	0.0 %	19825					2
7.5	7.500	7.409	-0.091	28691	0.0 %	28691					1
10.0	10.000	9.999	-0.001	38677	0.0 %	38677					0

LB134431 INMSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Jan 2025 14:05:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	26	69	-	0.0000	245.1	PPB	27 Jan 2025 14:14:59	S	Std
0.05 - 1	58.165	265	-	0.0500	245.1	PPB	27 Jan 2025 14:17:15	S	Std
0.2 - 1	80.3	728	-	0.2000	245.1	PPB	27 Jan 2025 14:19:35	S	Std
2.5 - 1	32.5	9976	-	2.5000	245.1	PPB	27 Jan 2025 14:21:52	S	Std
5.0 - 1	55	19825	-	5.0000	245.1	PPB	27 Jan 2025 14:24:08	S	Std
7.5 - 1	57.5	28691	-	7.5000	245.1	PPB	27 Jan 2025 14:26:25	S	Std
10.0 - 1	214	38677	-	10.0000	245.1	PPB	27 Jan 2025 14:28:43	S	Std
ICV69 - 1	ICV69	15548	4.0001	-245.1		PPB	27 Jan 2025 14:37:28	U	SMPL
ICB69 - 1	ICB69	60	-0.0167	-245.1		PPB	27 Jan 2025 14:39:43	U	SMPL
CCV71 - 1	CCV71	20001	5.1549	-245.1		PPB	27 Jan 2025 14:41:59	U	SMPL
CCB71 - 1	CCB71	63	-0.0159	-245.1		PPB	27 Jan 2025 14:44:14	U	SMPL
HighStd - 1	HighStd	38873	10.0493	-245.1		PPB	27 Jan 2025 14:48:46	U	SMPL
ChkStd - 1	ChkStd	25425	6.5616	-245.1		PPB	27 Jan 2025 14:51:02	U	SMPL
PB166288BL - 1	PBW	-163	-0.0746	-245.1		PPB	27 Jan 2025 14:53:20	U	SMPL
Q1066-02 - 1	EFF-WASTE WATER	90	-0.0089	-245.1		PPB	27 Jan 2025 14:57:54	U	SMPL
Q1066-02DUP - 1	EFF-WASTE WATERDUP	118	-0.0017	-245.1		PPB	27 Jan 2025 15:00:11	U	SMPL
Q1066-02MS - 1	EFF-WASTE WATERMS	12793	3.2856	-245.1		PPB	27 Jan 2025 15:05:06	U	SMPL
Q1066-02MSD - 1	EFF-WASTE WATERMSD	13018	3.3439	-245.1		PPB	27 Jan 2025 15:07:23	U	SMPL
CCV72 - 1	CCV72	19771	5.0953	-245.1		PPB	27 Jan 2025 15:09:40	U	SMPL
CCB72 - 1	CCB72	12	-0.0292	-245.1		PPB	27 Jan 2025 15:11:56	U	SMPL
Q1066-02LX5 - 1		-36	-0.0416	-245.1		PPB	27 Jan 2025 15:14:14	U	SMPL
Q1066-02A - 1		13977	3.5926	-245.1		PPB	27 Jan 2025 15:16:31	U	SMPL
CRA - 1	CRA	294	0.0440	-245.1		PPB	27 Jan 2025 15:27:37	U	SMPL
PB166288BS - 1	LCSW	14066	3.6157	-245.1		PPB	27 Jan 2025 15:32:26	U	SMPL
CCV73 - 1	CCV73	19707	5.0787	-245.1		PPB	27 Jan 2025 15:36:45	U	SMPL
CCB73 - 1	CCB73	111	-0.0035	-245.1		PPB	27 Jan 2025 15:39:00	U	SMPL