



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ardmore Chemical **SDG No.:** Q1066
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q1066 **SAS No.:** Q1066
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	6.15	6.0	102	40 - 160	P	01/20/2025	11:06	LB134350
	Copper	22.4	20.0	112	40 - 160	P	01/20/2025	11:06	LB134350
	Lead	12.2	12.0	102	40 - 160	P	01/20/2025	11:06	LB134350
	Nickel	40.3	40.0	101	40 - 160	P	01/20/2025	11:06	LB134350
	Zinc	43.0	40.0	108	40 - 160	P	01/20/2025	11:06	LB134350
CRA	Mercury	0.044	0.05	88	40 - 160	CV	01/27/2025	15:27	LB134431