



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Ardmore Chemical				SDG No.:	Q1066				
Contract:	ARDM01	Lab Code:	CHEM		Case No.:	Q1066	SAS No.:	Q1066		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB69	Mercury	0.20	+/-0.20	U	0.20	CV	01/27/2025	14:39	LB134431	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q1066

Contract: ARDM01

Lab Code: CHEM

Case No.: Q1066

SAS No.: Q1066

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB71	Mercury	0.20	+/-0.20	U	0.20	CV	01/27/2025	14:44	LB134431
CCB72	Mercury	0.20	+/-0.20	U	0.20	CV	01/27/2025	15:11	LB134431
CCB73	Mercury	0.20	+/-0.20	U	0.20	CV	01/27/2025	15:39	LB134431

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q1066

Contract: ARDM01

Lab Code: CHEM

Case No.: Q1066

SAS No.: Q1066

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	11:01	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	11:01	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	11:01	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	11:01	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	11:01	LB134350

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Ardmore Chemical</u>		SDG No.: <u>Q1066</u>							
Contract: <u>ARDM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1066</u>	SAS No.: <u>Q1066</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	11:36	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	11:36	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	11:36	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	11:36	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	11:36	LB134350
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	12:33	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	12:33	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	12:33	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	12:33	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	12:33	LB134350
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	13:39	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	13:39	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	13:39	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	13:39	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	13:39	LB134350
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	14:43	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	14:43	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	14:43	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	14:43	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	14:43	LB134350
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	15:41	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	15:41	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	15:41	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	15:41	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	15:41	LB134350
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	16:50	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	16:50	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	16:50	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	16:50	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	16:50	LB134350
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	17:37	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	17:37	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	17:37	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	17:37	LB134350
	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	17:37	LB134350
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	01/20/2025	17:56	LB134350
	Copper	20.0	+/-20.0	U	20.0	P	01/20/2025	17:56	LB134350
	Lead	12.0	+/-12.0	U	12.0	P	01/20/2025	17:56	LB134350
	Nickel	40.0	+/-40.0	U	40.0	P	01/20/2025	17:56	LB134350

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Ardmore Chemical</u>				SDG No.:	<u>Q1066</u>			
Contract:	<u>ARDM01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1066</u>	SAS No.:	<u>Q1066</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	01/20/2025	17:56	LB134350