

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
Lab Code: CHEM Case No.: Q1066 SAS No.: Q1066 SDG No.: Q1066
Instrument ID: BNA_F Calibration Date(s): 01/10/2025 01/10/2025
Calibration Time(s): 11:58 17:45

LAB FILE ID:		RRF2.5 = BF141109.D			RRF005 = BF141110.D		RRF010 = BF141111.D	
		RRF020 = BF141112.D			RRF040 = BF141113.D		RRF050 = BF141117.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
n-Nitrosodimethylamine		0.716	0.715	0.709	0.678	0.688	0.691	3.7
2-Fluorophenol		1.453	1.400	1.378	1.248	1.236	1.295	9.1
Phenol-d6		1.834	1.773	1.736	1.566	1.569	1.640	8.6
Phenol		1.943	1.902	1.861	1.701	1.665	1.756	8.4
bis(2-Chloroethyl) ether		1.445	1.372	1.363	1.256	1.279	1.308	6.9
2-Chlorophenol		1.561	1.476	1.458	1.345	1.338	1.389	8.2
2,2-oxybis(1-Chloropropane)		2.046	1.913	1.939	1.766	1.746	1.826	7.8
n-Nitroso-di-n-propylamine	1.062	1.068	1.011	1.009	0.906	0.915	0.966	8.4
Nitrobenzene-d5		0.400	0.390	0.394	0.365	0.370	0.377	4.8
Hexachloroethane		0.610	0.591	0.595	0.548	0.551	0.565	6.4
Nitrobenzene		0.419	0.410	0.411	0.377	0.384	0.393	5.2
Isophorone		0.692	0.668	0.674	0.617	0.629	0.645	5.1
2-Nitrophenol		0.161	0.173	0.190	0.178	0.184	0.179	5.3
2,4-Dimethylphenol		0.252	0.244	0.250	0.236	0.239	0.241	3.3
bis(2-Chloroethoxy) methane		0.435	0.419	0.430	0.387	0.393	0.405	5.6
2,4-Dichlorophenol		0.296	0.296	0.300	0.277	0.286	0.287	4.3
1,2,4-Trichlorobenzene		0.356	0.344	0.348	0.314	0.317	0.328	6.5
Naphthalene		1.178	1.126	1.131	0.998	1.023	1.055	8.6
Hexachlorobutadiene		0.213	0.207	0.208	0.188	0.194	0.197	6.2
4-Chloro-3-methylphenol		0.331	0.322	0.327	0.302	0.313	0.313	4.4
Hexachlorocyclopentadiene		0.170	0.181	0.201	0.200	0.189	0.188	5.7
2,4,6-Trichlorophenol		0.399	0.384	0.412	0.380	0.382	0.387	3.4
2-Fluorobiphenyl		1.582	1.459	1.408	1.225	1.224	1.310	13.4
2-Chloronaphthalene		1.345	1.288	1.273	1.164	1.174	1.210	7.8
Dimethylphthalate		1.447	1.406	1.404	1.299	1.318	1.343	5.8
Acenaphthylene		1.915	1.820	1.830	1.660	1.678	1.728	7.5
2,6-Dinitrotoluene		0.308	0.317	0.328	0.312	0.315	0.312	3.3
Acenaphthene		1.317	1.267	1.274	1.165	1.185	1.208	6.7
2,4-Dinitrophenol			0.097	0.133	0.155	0.168	0.148	19.4
4-Nitrophenol		0.224	0.241	0.255	0.250	0.264	0.248	5.3
2,4-Dinitrotoluene		0.372	0.401	0.423	0.395	0.424	0.402	4.7
Diethylphthalate		1.416	1.397	1.378	1.253	1.289	1.311	6.6
4-Chlorophenyl-phenylether		0.709	0.672	0.663	0.594	0.593	0.622	9.7
Fluorene		1.488	1.403	1.359	1.197	1.213	1.276	11.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
Lab Code: CHEM Case No.: Q1066 SAS No.: Q1066 SDG No.: Q1066
Instrument ID: BNA_F Calibration Date(s): 01/10/2025 01/10/2025
Calibration Time(s): 11:58 17:45

LAB FILE ID:		RRF2.5 = BF141109.D			RRF005 = BF141110.D		RRF010 = BF141111.D	
		RRF020 = BF141112.D			RRF040 = BF141113.D		RRF050 = BF141117.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
4,6-Dinitro-2-methylphenol		0.077	0.095	0.120	0.129	0.130	0.116	18.4
n-Nitrosodiphenylamine		0.706	0.677	0.683	0.651	0.603	0.645	7.4
Azobenzene		1.430	1.363	1.353	1.245	1.264	1.293	7.1
2,4,6-Tribromophenol		0.233	0.233	0.235	0.219	0.230	0.228	3.0
4-Bromophenyl-phenylether		0.241	0.237	0.238	0.232	0.220	0.230	4.2
Hexachlorobenzene		0.273	0.263	0.266	0.257	0.244	0.256	4.8
Pentachlorophenol		0.142	0.162	0.171	0.173	0.169	0.164	6.6
Phenanthrene		1.218	1.152	1.122	1.064	1.014	1.074	9.0
Anthracene		1.162	1.111	1.095	1.042	0.994	1.044	8.2
Di-n-butylphthalate		1.186	1.177	1.139	1.087	1.079	1.096	7.3
Fluoranthene		1.211	1.188	1.104	1.037	1.051	1.071	9.7
Benzidine		0.380	0.309	0.310	0.409	0.340	0.371	13.8
Pyrene		1.952	1.977	2.081	1.848	1.839	1.910	6.1
Terphenyl-d14		1.444	1.424	1.483	1.272	1.280	1.346	8.4
Butylbenzylphthalate		0.583	0.629	0.663	0.614	0.677	0.637	5.4
3,3-Dichlorobenzidine		0.434	0.422	0.444	0.418	0.426	0.433	2.7
Benzo (a) anthracene		1.473	1.420	1.427	1.245	1.329	1.361	6.3
Chrysene		1.382	1.316	1.336	1.207	1.237	1.274	5.5
Bis(2-ethylhexyl)phthalate		0.734	0.768	0.807	0.726	0.819	0.764	5.4
Di-n-octyl phthalate		0.950	1.039	1.169	1.151	1.235	1.154	10.5
Benzo (b) fluoranthene		1.461	1.415	1.290	1.201	1.315	1.290	9.0
Benzo (k) fluoranthene		1.128	1.102	1.141	1.043	1.038	1.090	3.7
Benzo (a) pyrene		1.103	1.090	1.095	1.039	1.048	1.064	3.3
Indeno (1,2,3-cd) pyrene		1.472	1.420	1.558	1.472	1.421	1.465	4.5
Dibenzo (a,h) anthracene		1.174	1.178	1.262	1.197	1.150	1.185	5.0
Benzo (g,h,i) perylene		1.256	1.221	1.347	1.263	1.195	1.233	7.6

All other compounds must meet a minimum RRF of 0.010.