

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF011025

Review By	yogesh	Review On	1/13/2025 12:24:27 AM
Supervise By	Jagrut	Supervise On	1/13/2025 11:38:11 AM
SubDirectory	BF011025	HP Acquire Method	BNA_F
		HP Processing Method	bf011025

STD. NAME	STD REF.#
Tune/Reschk	SP6701
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12647,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141108.D	10 Jan 2025 11:32		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141109.D	10 Jan 2025 11:58		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141110.D	10 Jan 2025 12:25	Compound #54 removed from 5ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141111.D	10 Jan 2025 13:17		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141112.D	10 Jan 2025 14:10		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141113.D	10 Jan 2025 14:36	The Calibration is Good For 8270 DOD and good for 625.1 Method	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141114.D	10 Jan 2025 15:03	Not Used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141115.D	10 Jan 2025 15:29	Not Used	RC/JU	Not Ok
9	SSTDICC080	SSTDICC080	BF141116.D	10 Jan 2025 15:56	Not Used	RC/JU	Not Ok
10	SSTDICC050	SSTDICC050	BF141117.D	10 Jan 2025 16:53		RC/JU	Ok,M
11	SSTDICC060	SSTDICC060	BF141118.D	10 Jan 2025 17:19		RC/JU	Ok
12	SSTDICC080	SSTDICC080	BF141119.D	10 Jan 2025 17:45	Compound #9 removed from 80ppm	RC/JU	Ok,M
13	SSTDICV040	ICVBF011025	BF141120.D	10 Jan 2025 18:12	Com# 69 & 77 failed in ICV for DOD	RC/JU	Ok
14	PB166008BL	PB166008BL	BF141121.D	10 Jan 2025 18:38		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF011625

Review By	Review On
Supervise By	Supervise On
SubDirectory	BF011625
HP Acquire Method	BNA_F
HP Processing Method	bf011025
STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12647,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141178.D	16 Jan 2025 10:25		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141179.D	16 Jan 2025 12:29		RC/JU	Ok
3	PB166074BL	PB166074BL	BF141180.D	16 Jan 2025 12:55		RC/JU	Ok
4	PB166074BS	PB166074BS	BF141181.D	16 Jan 2025 13:21		RC/JU	Ok,NS
5	PB166033BS	PB166033BS	BF141182.D	16 Jan 2025 13:48		RC/JU	Ok,NS
6	PB166033BSD	PB166033BSD	BF141183.D	16 Jan 2025 14:14		RC/JU	Ok,NS
7	PB166033BL	PB166033BL	BF141184.D	16 Jan 2025 14:40		RC/JU	Ok
8	Q1066-02	EFF-WASTE WATER	BF141185.D	16 Jan 2025 15:13	Surrogate Fail	RC/JU	Ok
9	Q1094-01	72-11938	BF141186.D	16 Jan 2025 15:39		RC/JU	Ok,NS
10	Q1094-01MS	72-11938MS	BF141187.D	16 Jan 2025 16:05		RC/JU	Ok,NS
11	Q1094-01MSD	72-11938MSD	BF141188.D	16 Jan 2025 16:32		RC/JU	Ok,NR
12	Q1098-01	OILY-STONE-COMP	BF141189.D	16 Jan 2025 16:58		RC/JU	Ok
13	Q1093-01	RT-3407	BF141190.D	16 Jan 2025 17:24	Internal Standard Fail	RC/JU	ReRun
14	Q1099-01	OR-02-01152025	BF141191.D	16 Jan 2025 17:50		RC/JU	Ok,NR

M : Manual Integration