

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: Q1066 SAS No.: Q1066 SDG No.: Q1066
 Instrument ID: MSVOA_N Calibration Date(s): 12/20/2024 12/20/2024
 Heated Purge: (Y/N) N Calibration Time(s): 10:05 11:42
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN085272.D		RRF020 = VN085273.D		RRF050 = VN085274.D			
		RRF100 = VN085275.D		RRF150 = VN085276.D		RRF =			
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	RRF	% RSD	
Chloromethane	2.717	2.623	2.537	2.356	2.430		2.533	5.7	
Vinyl Chloride	2.739	2.617	2.511	2.443	2.548		2.572	4.4	
Bromomethane	1.684	1.486	1.447	1.406	1.495		1.503	7.1	
Chloroethane	1.819	1.683	1.574	1.486	1.555		1.623	8	
Trichlorofluoromethane	3.660	3.639	3.515	3.359	3.492		3.533	3.5	
1,1-Dichloroethene	1.861	1.828	1.825	1.812	1.940		1.853	2.8	
Acrolein	0.327	0.221	0.221	0.212	0.255		0.247	19.4	
Acrylonitrile	0.962	1.074	1.078	1.080	1.158		1.070	6.5	
Methylene Chloride	2.390	2.246	2.145	2.069	2.193		2.208	5.4	
trans-1,2-Dichloroethene	1.886	1.924	1.933	1.885	2.011		1.928	2.7	
1,1-Dichloroethane	4.112	4.091	3.945	3.848	4.088		4.017	2.9	
Carbon Tetrachloride	0.643	0.605	0.581	0.573	0.561		0.593	5.5	
Chloroform	4.294	4.193	3.985	3.831	4.055		4.072	4.4	
1,1,1-Trichloroethane	0.743	0.695	0.657	0.649	0.641		0.677	6.3	
Benzene	1.746	1.700	1.610	1.606	1.570		1.646	4.5	
1,2-Dichloroethane	0.669	0.610	0.577	0.568	0.553		0.595	7.7	
Trichloroethene	0.383	0.368	0.354	0.353	0.352		0.362	3.7	
1,2-Dichloropropane	0.441	0.431	0.413	0.405	0.404		0.419	3.9	
Bromodichloromethane	0.645	0.633	0.595	0.592	0.587		0.610	4.4	
Toluene	1.976	1.904	1.792	1.773	1.770		1.843	5	
t-1,3-Dichloropropene	0.583	0.595	0.604	0.614	0.619		0.603	2.4	
cis-1,3-Dichloropropene	0.647	0.645	0.644	0.655	0.650		0.648	0.7	
1,1,2-Trichloroethane	0.421	0.402	0.367	0.365	0.355		0.382	7.4	
2-Chloroethyl vinyl ether	0.283	0.293	0.300	0.297	0.315		0.298	4	
Dibromochloromethane	0.475	0.466	0.437	0.431	0.429		0.448	4.8	
Tetrachloroethene	0.387	0.344	0.324	0.310	0.310		0.335	9.6	
Chlorobenzene	1.163	1.131	1.095	1.080	1.089		1.112	3.1	
Ethyl Benzene	1.810	1.922	1.957	2.001	2.007		1.939	4.1	
m/p-Xylenes	0.658	0.752	0.748	0.743	0.737		0.728	5.4	
o-Xylene	0.645	0.695	0.705	0.714	0.706		0.693	4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: Q1066 SAS No.: Q1066 SDG No.: Q1066
 Instrument ID: MSVOA_N Calibration Date(s): 12/20/2024 12/20/2024
 Heated Purge: (Y/N) N Calibration Time(s): 10:05 11:42
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN085272.D		RRF020 = VN085273.D		RRF050 = VN085274.D		
		RRF100 = VN085275.D		RRF150 = VN085276.D		RRF =		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	RRF	% RSD
Bromoform	0.304	0.294	0.284	0.291	0.288		0.292	2.6
1,1,2,2-Tetrachloroethane	0.698	0.626	0.594	0.586	0.583		0.617	7.8
1,3-Dichlorobenzene	0.775	0.818	0.808	0.810	0.798		0.802	2.1
1,4-Dichlorobenzene	0.741	0.754	0.794	0.791	0.798		0.776	3.4
1,2-Dichlorobenzene	0.730	0.768	0.784	0.766	0.760		0.762	2.6
1,2-Dichloroethane-d4	2.636	2.656	2.614	2.597	2.692		2.639	1.4
Toluene-d8	1.528	1.510	1.446	1.411	1.430		1.465	3.5
4-Bromofluorobenzene	0.468	0.488	0.484	0.489	0.501		0.486	2.5

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.