

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: Q1066 SAS No.: Q1066 SDG NO.: Q1066
 Lab File ID: VN085418.D Date Analyzed: 01/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
EFF-WASTE WATER	0	0.00				
VN0110WBL01	0	0.00				
VN0110WBS01	0	0.00				
VN0110WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.