



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Tetra Tech NUS, Inc.**

**Project Name: NWIRP Bethpage 112G08005-WE13**

**Project Manager: Ernie Wu**

**Chemtech Project # Q1069**

**Test Name: pH, Ignitability**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 01/09/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Flash Point, Ignitability, PCB, pH, SVOC-TCL BNA -20, TCLP Extraction, TCLP Mercury, TCLP Metal, TCLP VOA, TCLP ZHE Extraction and TCLP Metals Group1. This data package contains results for pH, Ignitability.

### **C. Analytical Techniques:**

The analysis of Ignitability was based on method 1030 and The analysis of pH was based on method 9045D.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for RW7B-CARBON-20250109 of pH as this sample received out of hold.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

**E. Additional Comments:** The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_