

## Initial and Continuing Calibration Verification

**Client:** Tetra Tech NUS, Inc.

**SDG No.:** Q1069

**Project:** NWIRP Bethpage 112G08005-WE13

**RunNo.:** LB134243

| Analyte                      | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>pH  | pH    | 7.02   | 7          | 100        | 90-110                 | 01/13/2025    |
| Sample ID: <b>CCV1</b><br>pH | pH    | 2.01   | 2.00       | 101        | 90-110                 | 01/13/2025    |
| Sample ID: <b>CCV2</b><br>pH | pH    | 12.02  | 12.00      | 100        | 90-110                 | 01/13/2025    |