


LB134281

7470A

INSTRUMENT ID: CY1

Linear



A= 0.0000e+000

B= 2.4141e-004

C= -1.3931e-001

Rho= 0.9990659

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |       |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.014 | 0.014  | 634   | 0.000      | 634   |       |       |       |       | 0/017 |
| 0.2    | 0.200  | 0.240 | 0.040  | 1571  | 0.0 %      | 1571  | 0     |       |       |       | 20    |
| 2.5    | 2.500  | 2.379 | -0.121 | 10432 | 0.0 %      | 10432 |       |       |       |       | -5    |
| 5.0    | 5.000  | 4.951 | -0.049 | 21086 | 0.0 %      | 21086 |       |       |       |       | 4     |
| 7.5    | 7.500  | 7.810 | 0.310  | 32929 | 0.0 %      | 32929 |       |       |       |       | 2     |
| 10.0   | 10.000 | 9.806 | -0.194 | 41197 | 0.0 %      | 41197 |       |       |       |       |       |

LB134281 INSTRUMENT ID : CV1

Method:7470A Operator : Mohan

Date Of Analysis : 14 january 2025 13:31:00

| Sample ID   | Extended ID          | $\mu$ Abs. | Conc.     | Std Conc | Method    | Units | Date            | Type |
|-------------|----------------------|------------|-----------|----------|-----------|-------|-----------------|------|
|             | 0                    | 634        | -         |          | 0 7470A   | PPB   | 1/14/2025 13:38 | Std  |
|             | 0.2                  | 1571       | -         |          | 0.2 7470A | PPB   | 1/14/2025 13:40 | Std  |
|             | 2.5                  | 10432      | -         |          | 2.5 7470A | PPB   | 1/14/2025 13:42 | Std  |
|             | 5                    | 21086      | -         |          | 5 7470A   | PPB   | 1/14/2025 13:45 | Std  |
|             | 7.5                  | 32929      | -         |          | 7.5 7470A | PPB   | 1/14/2025 13:52 | Std  |
|             | 10                   | 41197      | -         |          | 10 7470A  | PPB   | 1/14/2025 13:57 | Std  |
| ICV55       | ICV55                | 15842      | 3.6851 -  |          | 7470A     | PPB   | 1/14/2025 14:00 | SMPL |
| ICB55       | ICB55                | 412        | -0.0399 - |          | 7470A     | PPB   | 1/14/2025 14:02 | SMPL |
| CCV32       | CCV32                | 21245      | 4.9895 -  |          | 7470A     | PPB   | 1/14/2025 14:07 | SMPL |
| CCB32       | CCB32                | 494        | -0.0201 - |          | 7470A     | PPB   | 1/14/2025 14:09 | SMPL |
| CRA         | CRA                  | 1448       | 0.2103 -  |          | 7470A     | PPB   | 1/14/2025 14:14 | SMPL |
| HighStd     | HighStd              | 40086      | 9.5379 -  |          | 7470A     | PPB   | 1/14/2025 14:16 | SMPL |
| ChkStd      | ChkStd               | 29320      | 6.9388 -  |          | 7470A     | PPB   | 1/14/2025 14:19 | SMPL |
| PB166050BL  | PBW                  | 658        | 0.0195 -  |          | 7470A     | PPB   | 1/14/2025 14:21 | SMPL |
| PB166050BS  | LCSW                 | 16009      | 3.7254 -  |          | 7470A     | PPB   | 1/14/2025 14:23 | SMPL |
| Q1036-02    | NP-WS-002            | 912        | 0.0809 -  |          | 7470A     | PPB   | 1/14/2025 14:26 | SMPL |
| Q1036-02DUP | NP-WS-002DUP         | 848        | 0.0654 -  |          | 7470A     | PPB   | 1/14/2025 14:28 | SMPL |
| Q1036-02MS  | NP-WS-002MS          | 16513      | 3.8471 -  |          | 7470A     | PPB   | 1/14/2025 14:30 | SMPL |
| Q1036-02MSD | NP-WS-002MSD         | 15749      | 3.6627 -  |          | 7470A     | PPB   | 1/14/2025 14:32 | SMPL |
| PB166051BL  | PBW                  | 857        | 0.0676 -  |          | 7470A     | PPB   | 1/14/2025 14:35 | SMPL |
| CCV33       | CCV33                | 20915      | 4.9098 -  |          | 7470A     | PPB   | 1/14/2025 14:37 | SMPL |
| CCB33       | CCB33                | 371        | -0.0497 - |          | 7470A     | PPB   | 1/14/2025 14:39 | SMPL |
| PB166051BS  | LCSW                 | 15676      | 3.645 -   |          | 7470A     | PPB   | 1/14/2025 14:42 | SMPL |
| Q1065-01    | MOO-25-0002          | -234       | -0.1958 - |          | 7470A     | PPB   | 1/14/2025 14:44 | SMPL |
| Q1069-02    | RW7B-CARBON-20250109 | 965        | 0.0936 -  |          | 7470A     | PPB   | 1/14/2025 14:46 | SMPL |
| Q1074-02    | HACKENSACK-DGA       | 523        | -0.0131 - |          | 7470A     | PPB   | 1/14/2025 14:48 | SMPL |
| Q1076-02    | ARS20-0006           | 745        | 0.0405 -  |          | 7470A     | PPB   | 1/14/2025 14:58 | SMPL |
| Q1076-02DUP | ARS20-0006DUP        | 778        | 0.0485 -  |          | 7470A     | PPB   | 1/14/2025 15:00 | SMPL |
| Q1076-02MS  | ARS20-0006MS         | 17256      | 4.0265 -  |          | 7470A     | PPB   | 1/14/2025 15:02 | SMPL |
| Q1076-02MSD | ARS20-0006MSD        | 17222      | 4.0183 -  |          | 7470A     | PPB   | 1/14/2025 15:05 | SMPL |
| PB166025TB  | PB166025TB           | 737        | 0.0386 -  |          | 7470A     | PPB   | 1/14/2025 15:07 | SMPL |
| PB166026TB  | PB166026TB           | 931        | 0.0854 -  |          | 7470A     | PPB   | 1/14/2025 15:09 | SMPL |
| CCV34       | CCV34                | 20017      | 4.693 -   |          | 7470A     | PPB   | 1/14/2025 15:12 | SMPL |
| CCB34       | CCB34                | 441        | -0.0328 - |          | 7470A     | PPB   | 1/14/2025 15:14 | SMPL |
| Q1036-02LX5 |                      | 322        | -0.0616 - |          | 7470A     | PPB   | 1/14/2025 15:16 | SMPL |
| Q1036-02A   |                      | 14699      | 3.4092 -  |          | 7470A     | PPB   | 1/14/2025 15:18 | SMPL |
| Q1076-02LX5 |                      | 69         | -0.1227 - |          | 7470A     | PPB   | 1/14/2025 15:21 | SMPL |
| Q1076-02A   |                      | 16853      | 3.9292 -  |          | 7470A     | PPB   | 1/14/2025 15:23 | SMPL |
| CCV35       | CCV35                | 19447      | 4.5554 -  |          | 7470A     | PPB   | 1/14/2025 15:25 | SMPL |
| CCB35       | CCB35                | 340        | -0.0572 - |          | 7470A     | PPB   | 1/14/2025 15:28 | SMPL |