



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1069
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1069 **SAS No.:** Q1069
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	105	40 - 160	CV	01/14/2025	14:14	LB134281
CRI01	Arsenic	19.2	20.0	96	40 - 160	P	01/17/2025	11:35	LB134334
	Barium	85.3	100	85	40 - 160	P	01/17/2025	11:35	LB134334
	Cadmium	5.82	6.0	97	40 - 160	P	01/17/2025	11:35	LB134334
	Chromium	9.43	10.0	94	40 - 160	P	01/17/2025	11:35	LB134334
	Lead	12.4	12.0	104	40 - 160	P	01/17/2025	11:35	LB134334
	Selenium	20.7	20.0	104	40 - 160	P	01/17/2025	11:35	LB134334
	Silver	10.00	10.0	100	40 - 160	P	01/17/2025	11:35	LB134334