

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1069

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166026TB		WATER		Batch Number:		PB166051		Prep Date:	01/14/2025	
	Mercury	0.85	<2.00	J	1.60	2.00	CV	01/14/2025	15:09	LB134281
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166051BL		WATER		Batch Number:		PB166051		Prep Date:	01/14/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	01/14/2025	14:35	LB134281

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1069

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166026TB		WATER		Batch Number:		PB166049		Prep Date:	01/14/2025	
	Arsenic	100	<100	U	80.0	100	P	01/17/2025	12:42	LB134334
	Barium	500	<500	U	125	500	P	01/17/2025	12:42	LB134334
	Cadmium	30.0	<30.0	U	7.50	30.0	P	01/17/2025	12:42	LB134334
	Chromium	50.0	<50.0	U	25.0	50.0	P	01/17/2025	12:42	LB134334
	Lead	60.0	<60.0	U	48.0	60.0	P	01/17/2025	12:42	LB134334
	Selenium	100	<100	U	80.0	100	P	01/17/2025	12:42	LB134334
	Silver	50.0	<50.0	U	25.0	50.0	P	01/17/2025	12:42	LB134334
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166049BL		WATER		Batch Number:		PB166049		Prep Date:	01/14/2025	
	Arsenic	100	<100	U	80.0	100	P	01/17/2025	12:34	LB134334
	Barium	500	<500	U	125	500	P	01/17/2025	12:34	LB134334
	Cadmium	30.0	<30.0	U	7.50	30.0	P	01/17/2025	12:34	LB134334
	Chromium	50.0	<50.0	U	25.0	50.0	P	01/17/2025	12:34	LB134334
	Lead	60.0	<60.0	U	48.0	60.0	P	01/17/2025	12:34	LB134334
	Selenium	100	<100	U	80.0	100	P	01/17/2025	12:34	LB134334
	Silver	50.0	<50.0	U	25.0	50.0	P	01/17/2025	12:34	LB134334