

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1069
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1069 **SAS No.:** Q1069
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	1.85			-20	20	01/17/2025	11:39	LB134334
	Barium	-1.17	6.0	20	-94	106	01/17/2025	11:39	LB134334
	Cadmium	-1.59	1.0	159	-5	7	01/17/2025	11:39	LB134334
	Chromium	55.0	52.0	106	42	62	01/17/2025	11:39	LB134334
	Lead	9.67			-12	12	01/17/2025	11:39	LB134334
	Selenium	-13.6			-20	20	01/17/2025	11:39	LB134334
	Silver	-2.42			-10	10	01/17/2025	11:39	LB134334
ICSAB01	Arsenic	110	104	106	88.4	120	01/17/2025	11:43	LB134334
	Barium	455	537	85	437	637	01/17/2025	11:43	LB134334
	Cadmium	971	972	100	826	1120	01/17/2025	11:43	LB134334
	Chromium	560	542	103	460	624	01/17/2025	11:43	LB134334
	Lead	57.5	49.0	117	37	61	01/17/2025	11:43	LB134334
	Selenium	31.7	46.0	69	26	66	01/17/2025	11:43	LB134334
	Silver	216	201	108	170	232	01/17/2025	11:43	LB134334