

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1069

Contract: TETR06

Lab Code: CHEM

Case No.: Q1069

SAS No.: Q1069

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166049BS							
Arsenic	ug/L	4000	3860		96	87 - 113	P
Barium	ug/L	1000	891		89	88 - 113	P
Cadmium	ug/L	1000	935		94	88 - 113	P
Chromium	ug/L	2000	2000		100	90 - 113	P
Lead	ug/L	5000	4740		95	86 - 113	P
Selenium	ug/L	10000	9430		94	83 - 114	P
Silver	ug/L	380	375		99	84 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1069

Contract: TETR06

Lab Code: CHEM

Case No.: Q1069

SAS No.: Q1069

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166051BS Mercury	ug/L	4.0	3.65		91	82 - 119	CV