

Contract:	<u>TETR06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1069</u>	SAS No.:	<u>Q1069</u>	SDG NO.:	<u>Q1069</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>01/06/2025</u>	<u>01/07/2025</u>		
		Calibration Times:		<u>19:57</u>	<u>03:16</u>		

LAB FILE ID:		RT 1000 =	<u>PP068915.D</u>	RT 750 =	<u>PP068916.D</u>
RT 500 =	PP068917.D	RT 250 =	PP068918.D	RT 050 =	PP068919.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.53	10.52	10.53	10.53	10.53	10.53	10.43	10.63
Tetrachloro-m-xylene	4.67	4.67	4.68	4.67	4.67	4.67	4.57	4.77

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>TETR06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1069</u>	SAS No.:	<u>Q1069</u>	SDG NO.:	<u>Q1069</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>01/06/2025</u>	<u>01/07/2025</u>		
		Calibration Times:		<u>19:57</u>	<u>03:16</u>		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP068915.D RT 750 = PP068916.D
RT 500 = PP068917.D RT 250 = PP068918.D RT 050 = PP068919.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	9.11	9.11	9.11	9.11	9.11	9.11	9.01	9.21
Tetrachloro-m-xylene	3.98	3.99	3.98	3.99	3.99	3.99	3.89	4.09