

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/13/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 09:16 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.83	5.84	5.74	5.94	0.01
Aroclor-1248-2 (2)	6.10	6.11	6.01	6.21	0.01
Aroclor-1248-3 (3)	6.31	6.31	6.21	6.41	0.01
Aroclor-1248-4 (4)	6.71	6.71	6.61	6.81	0.01
Aroclor-1248-5 (5)	6.75	6.75	6.65	6.85	0.01
Tetrachloro-m-xylene	4.67	4.67	4.57	4.77	0.00
Decachlorobiphenyl	10.51	10.53	10.43	10.63	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/13/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 09:16 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.08	5.09	4.99	5.19	0.01
Aroclor-1248-2	(2)	5.32	5.32	5.22	5.42	0.00
Aroclor-1248-3	(3)	5.36	5.37	5.27	5.47	0.01
Aroclor-1248-4	(4)	5.54	5.54	5.44	5.64	0.00
Aroclor-1248-5	(5)	5.93	5.94	5.84	6.04	0.01
Tetrachloro-m-xylene		3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl		9.10	9.11	9.01	9.21	0.01