

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/13/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 15:26 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.68	6.69	6.59	6.79	0.01
Aroclor-1254-2	(2)	6.90	6.91	6.81	7.01	0.01
Aroclor-1254-3	(3)	7.27	7.27	7.17	7.37	0.01
Aroclor-1254-4	(4)	7.55	7.55	7.45	7.65	0.00
Aroclor-1254-5	(5)	7.97	7.97	7.87	8.07	0.00
Tetrachloro-m-xylene		4.67	4.67	4.57	4.77	0.00
Decachlorobiphenyl		10.52	10.52	10.42	10.62	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/13/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 15:26 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.89	5.90	5.80	6.00	0.01
Aroclor-1254-2 (2)	6.04	6.05	5.95	6.15	0.01
Aroclor-1254-3 (3)	6.45	6.45	6.35	6.55	0.00
Aroclor-1254-4 (4)	6.68	6.68	6.58	6.78	0.00
Aroclor-1254-5 (5)	7.10	7.10	7.00	7.20	0.00
Tetrachloro-m-xylene	3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl	9.11	9.11	9.01	9.21	0.00