

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/13/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 19:19 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-2	(2)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-3	(3)	5.91	5.92	5.82	6.02	0.01
Aroclor-1016-4	(4)	6.01	6.02	5.92	6.12	0.01
Aroclor-1016-5	(5)	6.31	6.31	6.21	6.41	0.01
Aroclor-1260-1	(1)	7.43	7.43	7.33	7.53	0.00
Aroclor-1260-2	(2)	7.68	7.69	7.59	7.79	0.01
Aroclor-1260-3	(3)	8.04	8.05	7.95	8.15	0.01
Aroclor-1260-4	(4)	8.27	8.28	8.18	8.38	0.01
Aroclor-1260-5	(5)	8.61	8.61	8.51	8.71	0.00
Tetrachloro-m-xylene		4.67	4.67	4.57	4.77	0.00
Decachlorobiphenyl		10.51	10.52	10.42	10.62	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/13/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 19:19 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.08	5.09	4.99	5.19	0.01
Aroclor-1016-2	(2)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-3	(3)	5.28	5.28	5.18	5.38	0.00
Aroclor-1016-4	(4)	5.32	5.32	5.22	5.42	0.00
Aroclor-1016-5	(5)	5.54	5.54	5.44	5.64	0.00
Aroclor-1260-1	(1)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-2	(2)	6.77	6.77	6.67	6.87	0.00
Aroclor-1260-3	(3)	6.92	6.93	6.83	7.03	0.01
Aroclor-1260-4	(4)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-5	(5)	7.64	7.64	7.54	7.74	0.00
Tetrachloro-m-xylene		3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl		9.10	9.11	9.01	9.21	0.01