

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/14/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 08:53 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.83	5.83	5.73	5.93	0.00
Aroclor-1242-2 (2)	5.85	5.86	5.76	5.96	0.01
Aroclor-1242-3 (3)	5.91	5.92	5.82	6.02	0.01
Aroclor-1242-4 (4)	6.01	6.02	5.92	6.12	0.01
Aroclor-1242-5 (5)	6.75	6.75	6.65	6.85	0.01
Tetrachloro-m-xylene	4.67	4.67	4.57	4.77	0.00
Decachlorobiphenyl	10.51	10.53	10.43	10.63	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

Continuing Calib Date: 01/14/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 08:53 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.08	5.08	4.98	5.18	0.00
Aroclor-1242-2	(2)	5.10	5.10	5.00	5.20	0.00
Aroclor-1242-3	(3)	5.28	5.28	5.18	5.38	0.00
Aroclor-1242-4	(4)	5.36	5.37	5.27	5.47	0.01
Aroclor-1242-5	(5)	5.89	5.90	5.80	6.00	0.01
Tetrachloro-m-xylene		3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl		9.10	9.11	9.01	9.21	0.01