

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL01 Date Analyzed: 01/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069070.D Time Analyzed: 08:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.829	5.732	5.932	457.050	500.000	-8.6
Aroclor-1016-2	5.852	5.754	5.954	466.230	500.000	-6.8
Aroclor-1016-3	5.914	5.818	6.018	444.130	500.000	-11.2
Aroclor-1016-4	6.012	5.916	6.116	443.710	500.000	-11.3
Aroclor-1016-5	6.307	6.210	6.410	443.670	500.000	-11.3
Aroclor-1260-1	7.429	7.333	7.533	429.220	500.000	-14.2
Aroclor-1260-2	7.682	7.587	7.787	447.600	500.000	-10.5
Aroclor-1260-3	8.042	7.947	8.147	439.650	500.000	-12.1
Aroclor-1260-4	8.275	8.180	8.380	433.410	500.000	-13.3
Aroclor-1260-5	8.606	8.513	8.713	458.940	500.000	-8.2
Decachlorobiphenyl	10.515	10.424	10.624	42.770	50.000	-14.5
Tetrachloro-m-xylene	4.669	4.572	4.772	48.130	50.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL01 Date Analyzed: 01/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069070.D Time Analyzed: 08:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.080	4.985	5.185	490.400	500.000	-1.9
Aroclor-1016-2	5.100	5.004	5.204	470.090	500.000	-6.0
Aroclor-1016-3	5.279	5.184	5.384	491.300	500.000	-1.7
Aroclor-1016-4	5.319	5.224	5.424	472.040	500.000	-5.6
Aroclor-1016-5	5.537	5.442	5.642	471.610	500.000	-5.7
Aroclor-1260-1	6.578	6.484	6.684	444.560	500.000	-11.1
Aroclor-1260-2	6.765	6.672	6.872	454.750	500.000	-9.1
Aroclor-1260-3	6.921	6.828	7.028	439.100	500.000	-12.2
Aroclor-1260-4	7.395	7.302	7.502	434.160	500.000	-13.2
Aroclor-1260-5	7.634	7.541	7.741	444.140	500.000	-11.2
Decachlorobiphenyl	9.101	9.011	9.211	42.450	50.000	-15.1
Tetrachloro-m-xylene	3.980	3.883	4.083	49.630	50.000	-0.7