

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL02 Date Analyzed: 01/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069084.D Time Analyzed: 14:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.831	5.732	5.932	462.450	500.000	-7.5
Aroclor-1016-2	5.853	5.754	5.954	468.150	500.000	-6.4
Aroclor-1016-3	5.916	5.818	6.018	446.190	500.000	-10.8
Aroclor-1016-4	6.014	5.916	6.116	447.590	500.000	-10.5
Aroclor-1016-5	6.307	6.210	6.410	446.160	500.000	-10.8
Aroclor-1260-1	7.430	7.333	7.533	435.620	500.000	-12.9
Aroclor-1260-2	7.684	7.587	7.787	452.190	500.000	-9.6
Aroclor-1260-3	8.044	7.947	8.147	441.540	500.000	-11.7
Aroclor-1260-4	8.276	8.180	8.380	441.260	500.000	-11.7
Aroclor-1260-5	8.609	8.513	8.713	459.450	500.000	-8.1
Decachlorobiphenyl	10.517	10.424	10.624	43.490	50.000	-13.0
Tetrachloro-m-xylene	4.670	4.572	4.772	48.350	50.000	-3.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL02 Date Analyzed: 01/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069084.D Time Analyzed: 14:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.083	4.985	5.185	482.760	500.000	-3.4
Aroclor-1016-2	5.102	5.004	5.204	466.190	500.000	-6.8
Aroclor-1016-3	5.282	5.184	5.384	479.590	500.000	-4.1
Aroclor-1016-4	5.321	5.224	5.424	459.170	500.000	-8.2
Aroclor-1016-5	5.539	5.442	5.642	460.090	500.000	-8.0
Aroclor-1260-1	6.580	6.484	6.684	443.550	500.000	-11.3
Aroclor-1260-2	6.768	6.672	6.872	457.970	500.000	-8.4
Aroclor-1260-3	6.924	6.828	7.028	439.710	500.000	-12.1
Aroclor-1260-4	7.399	7.302	7.502	430.540	500.000	-13.9
Aroclor-1260-5	7.637	7.541	7.741	439.440	500.000	-12.1
Decachlorobiphenyl	9.105	9.011	9.211	42.960	50.000	-14.1
Tetrachloro-m-xylene	3.982	3.883	4.083	48.550	50.000	-2.9