

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL03 Date Analyzed: 01/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP069104.D Time Analyzed: 08:37

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.830  | 5.732     | 5.932  | 467.670            | 500.000           | -6.5  |
| Aroclor-1016-2       | 5.852  | 5.754     | 5.954  | 487.180            | 500.000           | -2.6  |
| Aroclor-1016-3       | 5.915  | 5.818     | 6.018  | 472.270            | 500.000           | -5.5  |
| Aroclor-1016-4       | 6.013  | 5.916     | 6.116  | 469.660            | 500.000           | -6.1  |
| Aroclor-1016-5       | 6.307  | 6.210     | 6.410  | 469.370            | 500.000           | -6.1  |
| Aroclor-1260-1       | 7.430  | 7.333     | 7.533  | 450.250            | 500.000           | -10.0 |
| Aroclor-1260-2       | 7.684  | 7.587     | 7.787  | 466.880            | 500.000           | -6.6  |
| Aroclor-1260-3       | 8.043  | 7.947     | 8.147  | 457.860            | 500.000           | -8.4  |
| Aroclor-1260-4       | 8.276  | 8.180     | 8.380  | 455.180            | 500.000           | -9.0  |
| Aroclor-1260-5       | 8.607  | 8.513     | 8.713  | 474.780            | 500.000           | -5.0  |
| Decachlorobiphenyl   | 10.515 | 10.424    | 10.624 | 44.680             | 50.000            | -10.6 |
| Tetrachloro-m-xylene | 4.670  | 4.572     | 4.772  | 49.610             | 50.000            | -0.8  |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL03 Date Analyzed: 01/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP069104.D Time Analyzed: 08:37

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 5.080 | 4.985     | 5.185 | 503.810            | 500.000           | 0.8   |
| Aroclor-1016-2       | 5.099 | 5.004     | 5.204 | 487.740            | 500.000           | -2.5  |
| Aroclor-1016-3       | 5.278 | 5.184     | 5.384 | 504.930            | 500.000           | 1.0   |
| Aroclor-1016-4       | 5.318 | 5.224     | 5.424 | 484.070            | 500.000           | -3.2  |
| Aroclor-1016-5       | 5.536 | 5.442     | 5.642 | 492.090            | 500.000           | -1.6  |
| Aroclor-1260-1       | 6.577 | 6.484     | 6.684 | 455.070            | 500.000           | -9.0  |
| Aroclor-1260-2       | 6.765 | 6.672     | 6.872 | 463.400            | 500.000           | -7.3  |
| Aroclor-1260-3       | 6.921 | 6.828     | 7.028 | 447.620            | 500.000           | -10.5 |
| Aroclor-1260-4       | 7.395 | 7.302     | 7.502 | 453.820            | 500.000           | -9.2  |
| Aroclor-1260-5       | 7.634 | 7.541     | 7.741 | 463.570            | 500.000           | -7.3  |
| Decachlorobiphenyl   | 9.101 | 9.011     | 9.211 | 46.610             | 50.000            | -6.8  |
| Tetrachloro-m-xylene | 3.980 | 3.883     | 4.083 | 48.980             | 50.000            | -2.0  |