

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF011025

Review By	yogesh	Review On	1/13/2025 12:24:27 AM
Supervise By	Jagrut	Supervise On	1/13/2025 11:38:11 AM
SubDirectory	BF011025	HP Acquire Method	BNA_F
		HP Processing Method	bf011025
STD. NAME	STD REF.#		
Tune/Reschk	SP6701		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12647,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF141108.D	10 Jan 2025 11:32	RC/JU	Ok
2	SSTDICC2.5	BF141109.D	10 Jan 2025 11:58	RC/JU	Ok
3	SSTDICC005	BF141110.D	10 Jan 2025 12:25	RC/JU	Ok
4	SSTDICC010	BF141111.D	10 Jan 2025 13:17	RC/JU	Ok
5	SSTDICC020	BF141112.D	10 Jan 2025 14:10	RC/JU	Ok
6	SSTDICCC040	BF141113.D	10 Jan 2025 14:36	RC/JU	Ok
7	SSTDICC050	BF141114.D	10 Jan 2025 15:03	RC/JU	Not Ok
8	SSTDICC060	BF141115.D	10 Jan 2025 15:29	RC/JU	Not Ok
9	SSTDICC080	BF141116.D	10 Jan 2025 15:56	RC/JU	Not Ok
10	SSTDICC050	BF141117.D	10 Jan 2025 16:53	RC/JU	Ok,M
11	SSTDICC060	BF141118.D	10 Jan 2025 17:19	RC/JU	Ok
12	SSTDICC080	BF141119.D	10 Jan 2025 17:45	RC/JU	Ok,M
13	SSTDICV040	BF141120.D	10 Jan 2025 18:12	RC/JU	Ok
14	PB166008BL	BF141121.D	10 Jan 2025 18:38	RC/JU	Ok

M : Manual Integration