

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF011525

Review By	Jagrut	Review On	1/15/2025 5:55:24 PM
Supervise By	mohammad	Supervise On	1/16/2025 7:51:52 AM
SubDirectory	BF011525	HP Acquire Method	BNA_F
		HP Processing Method	bf011025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673		
CCC	SP6676		
Internal Standard/PEM	S12647,10ul/1000ul sample		
ICV/I.BLK	SP6686		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF141162.D	15 Jan 2025 09:08	RC/JU	Ok
2	SSTDCCC040	BF141163.D	15 Jan 2025 09:34	RC/JU	Ok
3	PB166052BL	BF141164.D	15 Jan 2025 10:00	RC/JU	Ok
4	PB166035BS	BF141165.D	15 Jan 2025 10:26	RC/JU	Ok,M
5	PB166026TB	BF141166.D	15 Jan 2025 10:52	RC/JU	Ok
6	PB166008BS	BF141167.D	15 Jan 2025 11:19	RC/JU	Ok,M
7	PB166008BSD	BF141168.D	15 Jan 2025 11:45	RC/JU	Ok,M
8	PB166035BL	BF141169.D	15 Jan 2025 12:11	RC/JU	Ok
9	PB166052BS	BF141170.D	15 Jan 2025 12:38	RC/JU	Ok,M
10	PB166061BL	BF141171.D	15 Jan 2025 13:12	RC/JU	Ok
11	PB166061BS	BF141172.D	15 Jan 2025 13:38	RC/JU	Ok,M
12	Q1069-01	BF141173.D	15 Jan 2025 14:13	RC/JU	ReRun
13	Q1069-01RX	BF141174.D	15 Jan 2025 14:39	RC/JU	Confirms
14	Q1068-01MS	BF141175.D	15 Jan 2025 15:05	RC/JU	Ok,M
15	Q1068-01MSD	BF141176.D	15 Jan 2025 16:40	RC/JU	Ok,M
16	SSTDCCC040	BF141177.D	15 Jan 2025 17:10	RC/JU	Ok

M : Manual Integration