

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF011525

SequenceID : BF011525

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. CCC fail high side for compound #77 which is not present in parameter list of the associated samples.	_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

_____ ✓ _____

a. B/N Fraction

d. Acid Fraction Surrogate recoveries failed in the original sample Q1069-01 and in the re-extracted sample Q1069-01RX, confirming the matrix interference.

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

_____ ✓ _____

a. B/N Fraction Recoveries of the compounds in Q1068-01MS and Q1068-01MSD were affected due to matrix interference and no corrective action is needed.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

_____ ✓ _____

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

_____ ✓ _____

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

_____ ✓ _____

ADDITIONAL COMMENTS:

Recovery fail for some compound in PB166035BS, PB166008BS/BSD, PB166052BS and PB166061BS which hit are not present in any associated samples.

Jagrut
Analyst

01/15/2025
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF011525

SequenceID : BF011525 NA NO YES