

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF011025**

Review By	yogesh	Review On	1/13/2025 12:24:27 AM		
Supervise By	Jagrut	Supervise On	1/13/2025 11:38:11 AM		
SubDirectory	BF011025	HP Acquire Method	BNA_F	HP Processing Method	bf011025

STD. NAME	STD REF.#
Tune/Reschk	SP6701
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12647,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141108.D	10 Jan 2025 11:32		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141109.D	10 Jan 2025 11:58		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141110.D	10 Jan 2025 12:25	Compound #54 removed from 5ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141111.D	10 Jan 2025 13:17		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141112.D	10 Jan 2025 14:10		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141113.D	10 Jan 2025 14:36	The Calibration is Good For 8270 DOD and good for 625.1 Method	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141114.D	10 Jan 2025 15:03	Not Used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141115.D	10 Jan 2025 15:29	Not Used	RC/JU	Not Ok
9	SSTDICC080	SSTDICC080	BF141116.D	10 Jan 2025 15:56	Not Used	RC/JU	Not Ok
10	SSTDICC050	SSTDICC050	BF141117.D	10 Jan 2025 16:53		RC/JU	Ok,M
11	SSTDICC060	SSTDICC060	BF141118.D	10 Jan 2025 17:19		RC/JU	Ok
12	SSTDICC080	SSTDICC080	BF141119.D	10 Jan 2025 17:45	Compound #9 removed from 80ppm	RC/JU	Ok,M
13	SSTDICV040	ICVBF011025	BF141120.D	10 Jan 2025 18:12	Com# 69 & 77 failed in ICV for DOD	RC/JU	Ok
14	PB166008BL	PB166008BL	BF141121.D	10 Jan 2025 18:38		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF011525**

Review By	Jagrut	Review On	1/15/2025 5:55:24 PM		
Supervise By	mohammad	Supervise On	1/16/2025 7:51:52 AM		
SubDirectory	BF011525	HP Acquire Method	BNA_F	HP Processing Method	bf011025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12647,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141162.D	15 Jan 2025 09:08		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141163.D	15 Jan 2025 09:34		RC/JU	Ok
3	PB166052BL	PB166052BL	BF141164.D	15 Jan 2025 10:00		RC/JU	Ok
4	PB166035BS	PB166035BS	BF141165.D	15 Jan 2025 10:26		RC/JU	Ok,M
5	PB166026TB	PB166026TB	BF141166.D	15 Jan 2025 10:52		RC/JU	Ok
6	PB166008BS	PB166008BS	BF141167.D	15 Jan 2025 11:19		RC/JU	Ok,M
7	PB166008BSD	PB166008BSD	BF141168.D	15 Jan 2025 11:45		RC/JU	Ok,M
8	PB166035BL	PB166035BL	BF141169.D	15 Jan 2025 12:11		RC/JU	Ok
9	PB166052BS	PB166052BS	BF141170.D	15 Jan 2025 12:38		RC/JU	Ok,M
10	PB166061BL	PB166061BL	BF141171.D	15 Jan 2025 13:12		RC/JU	Ok
11	PB166061BS	PB166061BS	BF141172.D	15 Jan 2025 13:38		RC/JU	Ok,M
12	Q1069-01	RW7B-CARBON-20250	BF141173.D	15 Jan 2025 14:13	Surrogate Fail - RE EXTRACTION.	RC/JU	ReRun
13	Q1069-01RX	RW7B-CARBON-20250	BF141174.D	15 Jan 2025 14:39	Surrogate Fail	RC/JU	Confirms
14	Q1068-01MS	TR-06-1-10-2025MS	BF141175.D	15 Jan 2025 15:05		RC/JU	Ok,M
15	Q1068-01MSD	TR-06-1-10-2025MSD	BF141176.D	15 Jan 2025 16:40		RC/JU	Ok,M
16	SSTDCCC040	SSTDCCC040EC	BF141177.D	15 Jan 2025 17:10		RC/JU	Ok

M : Manual Integration