

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1069 SDG NO.: Q1069

Lab File ID: BF141108.D

DFTPP Injection Date: 01/10/2025

Instrument ID: BNA_F

DFTPP Injection Time: 11:32

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.2
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	38
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	49.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.9
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	13.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.1 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF141109.D	01/10/2025	11:58
SSTDICC005	SSTDICC005	BF141110.D	01/10/2025	12:25
SSTDICC010	SSTDICC010	BF141111.D	01/10/2025	13:17
SSTDICC020	SSTDICC020	BF141112.D	01/10/2025	14:10
SSTDICCC040	SSTDICCC040	BF141113.D	01/10/2025	14:36
SSTDICC050	SSTDICC050	BF141117.D	01/10/2025	16:53
SSTDICC060	SSTDICC060	BF141118.D	01/10/2025	17:19
SSTDICC080	SSTDICC080	BF141119.D	01/10/2025	17:45