

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG No.: Q1069

Instrument ID: MSVOA_N Calibration Date/Time: 01/16/2025 12:30

Lab File ID: VN085465.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.737	0.708		-3.93	20
1,1-Dichloroethene	0.537	0.539		0.37	20
2-Butanone	0.384	0.397		3.38	20
Carbon Tetrachloride	0.557	0.567		1.79	20
Chloroform	1.218	1.244		2.13	20
Benzene	1.463	1.531		4.65	20
1,2-Dichloroethane	0.551	0.573		3.99	20
Trichloroethene	0.341	0.331		-2.93	20
Tetrachloroethene	0.341	0.346		1.47	20
Chlorobenzene	1.095	1.103	0.3	0.73	20
1,2-Dichloroethane-d4	0.807	0.883		9.42	20
Dibromofluoromethane	0.347	0.381		9.8	20
Toluene-d8	1.232	1.408		14.29	20
4-Bromofluorobenzene	0.422	0.495		17.3	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.