

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG No.: Q1069

Instrument ID: MSVOA_N Calibration Date/Time: 01/16/2025 18:32

Lab File ID: VN085478.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.737	0.742		0.68	50
1,1-Dichloroethene	0.537	0.556		3.54	50
2-Butanone	0.384	0.428		11.46	50
Carbon Tetrachloride	0.557	0.585		5.03	50
Chloroform	1.218	1.273		4.52	50
Benzene	1.463	1.568		7.18	50
1,2-Dichloroethane	0.551	0.592		7.44	50
Trichloroethene	0.341	0.348		2.05	50
Tetrachloroethene	0.341	0.352		3.23	50
Chlorobenzene	1.095	1.158	0.3	5.75	50
1,2-Dichloroethane-d4	0.807	0.859		6.44	50
Dibromofluoromethane	0.347	0.371		6.92	50
Toluene-d8	1.232	1.344		9.09	50
4-Bromofluorobenzene	0.422	0.476		12.8	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.