

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1069
Client: Tetra Tech NUS, Inc.
Analytical Method: SW8260-Low **Datafile :** VN085477.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0116WBSD02	Vinyl chloride	20	19.8	ug/L	99	12		58	137	20
	1,1-Dichloroethene	20	20.4	ug/L	102	10		71	131	20
	2-Butanone	100	110	ug/L	110	23	*	56	143	20
	Carbon Tetrachloride	20	20.4	ug/L	102	10		72	136	20
	Chloroform	20	20.7	ug/L	104	14		79	124	20
	Benzene	20	20.5	ug/L	103	9		79	120	20
	1,2-Dichloroethane	20	20.7	ug/L	104	10		73	128	20
	Trichloroethene	20	19.7	ug/L	99	7		79	123	20
	Tetrachloroethene	20	20.5	ug/L	103	5		74	129	20
	Chlorobenzene	20	20.1	ug/L	101	5		82	118	20