

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETRO6
 Lab Code: CHEM Case No.: Q1069 SAS No.: Q1069 SDG NO.: Q1069
 Lab File ID: VN085465.D Date Analyzed: 01/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	199641	8.22	338720	9.09	304435	11.87
UPPER LIMIT	399282	8.724	677440	9.594	608870	12.365
LOWER LIMIT	99820.5	7.724	169360	8.594	152218	11.365
EPA SAMPLE NO.						
RW7B-CARBON-20250109	198665	8.22	358567	9.09	314710	11.86
VN0116WBL01	181190	8.22	338995	9.10	304427	11.87
VN0116WBS02	216168	8.22	362531	9.10	315947	11.87
VN0116WBSD02	189081	8.22	325305	9.10	284195	11.86

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.