

Cover Page

Order ID : Q1070

Project ID : W.T.P. Final Effluent 2025

Client : Mars Chocolate North America, LLC

Lab Sample Number

Q1070-01
Q1070-02
Q1070-03
Q1070-04

Client Sample Number

1 EFFLUENT COMPOSITE
2 EFFLUENT GRAB
Q1070-02MS
Q1070-02MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 1/17/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Mars Chocolate North America, LLC

Project Name: W.T.P. Final Effluent 2025

Project # N/A

Chemtech Project # Q1070

Test Name: Oil and Grease,Color,Ammonia,BOD Soluble,Phosphorus-Total,TDS,COD,BOD5,TSS

A. Number of Samples and Date of Receipt:

4 Water samples were received on 01/10/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, BOD Soluble, BOD5, COD, Color, Oil and Grease, Phosphorus-Total, TDS and TSS. This data package contains results for Oil and Grease,Color,Ammonia,BOD Soluble,Phosphorus-Total,TDS,COD,BOD5,TSS.

C. Analytical Techniques:

The analysis of Oil and Grease was based on method 1664A, The analysis of Phosphorus-Total was based on method 365.3, The analysis of Color was based on method SM2120 B, The analysis of TDS was based on method SM2540 C, The analysis of TSS was based on method SM2540 D, The analysis of Ammonia was based on method SM4500-NH3, The analysis of BOD Soluble,BOD5 was based on method SM5210 B and The analysis of COD was based on method SM5220 D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample 1 EFFLUENT COMPOSITE was diluted due to high concentrations for Phosphorus, Total.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Q1070-01 sample was analyzed with 20X straight Dilutions for Color, Due to brown matrix.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1070

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: KETAN PATEL

Date: 01/17/2025