

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1070

Project: W.T.P. Final Effluent 2025

RunNo.: LB134293

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.482	0.50	96	90-110	01/15/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.488	0.50	98	90-110	01/15/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.482	0.50	96	90-110	01/15/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.488	0.50	98	90-110	01/15/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1070

Project: W.T.P. Final Effluent 2025

RunNo.: LB134305

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1070

Project: W.T.P. Final Effluent 2025

RunNo.: LB134308

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: CCV1 COD	mg/L	51.341	50	103	95-105	01/16/2025
Sample ID: CCV2 COD	mg/L	50.336	50	101	95-105	01/16/2025
Sample ID: ICV COD	mg/L	51.341	50	103	95-105	10/14/2024