
From: Hairston, Miles (NE) <Miles.Hairston@gdit.com>
Sent: Monday, January 13, 2025 2:04 PM
To: Deepak Parmar
Cc: Sohil Jodhani; Mohammad Ahmed
Subject: Region 07 | Case 51376 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC
Attachments: SKM_95825011309260.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Please note that MA 3107.2 is not longer needed for the water TVOA samples of Case 51376. The vendor comments have been updated to include that information.

Thanks,
Miles Hairston
Associate Environmental Analyst
Under contract to EPA
QSS Coordinator – EPA Regions 7, 8, and 9

Work Phone: +1 571-454-0346
Miles.Hairston@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

Leave alert: 1/9/2025

GENERAL DYNAMICS
a Hamilton Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Monday, January 13, 2025 9:49 AM
To: Hairston, Miles (NE) <Miles.Hairston@gdit.com>
Cc: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Subject: Region 07 | Case 51376 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

As per ASR MA number mention on comment however, received COC is not mentioned any MA . Therefor Lab like to know how to process ?

Please see attachment for your reference.

Thanks & Regards,



Deepak Parmar
QA/QC
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3154
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com



From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Thursday, January 23, 2025 2:26 PM
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: Evans, Barry
Subject: TASK AREA DAS | REGION 7 | CASE 51376 | LAB ACE | SDG GCQ67 & GCQ85 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: Q1071-02(GCQ68).pdf; Q1071-03(GCQ69).pdf; Q1071-04(GCQ70).pdf; Q1071-05(GCQ71).pdf; Q1071-06(GCQ72).pdf; Q1071-07(GCQ73).pdf; Q1071-08(GCQ74).pdf; Q1071-09(GCQ75).pdf; Q1071-10(GCQ76).pdf; Q1071-14(GCQ78).pdf; Q1071-15(GCQ79).pdf; Q1071-16(GCQ80).pdf; Q1072-16(GCQ98).pdf; Q1072-15(GCQ97).pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Please refer to the Region's response below regarding Case 51376/SDGs GCQ67 & GCQ85. Also, please be sure to provide detailed information about this issue in the SDG Narrative and let me know if you have any questions.

ISSUE: "Lab is sending this email with regards to case 51376.

Lab has received water samples for TVOA analysis. Lab has analyzed samples for TVOA analysis for the samples in a continuous analytical sequence where samples are having high concentrations of target analytes and required dilution as well as you can review attached quant reports for the samples. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables. Please note that there is no any additional Lab QC failure associated to undiluted and dilution analysis for the samples as per attached.

Please see attached."

REGION: "The labs approach outlined below is acceptable and they should proceed accordingly. Please have them thoroughly document the occurrences in the case narrative."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Evans, Barry <Evans.Barry@epa.gov>

Sent: Thursday, January 23, 2025 1:27 PM

To: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Subject: RE: TASK AREA DAS | REGION 7 | CASE 51376 | LAB ACE | SDG GCQ67 & GCQ85 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT |

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good Afternoon,

The labs approach outlined below is acceptable and they should proceed accordingly. Please have them thoroughly document the occurrences in the case narrative.

Thanks,
Barry

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Sent: Thursday, January 23, 2025 12:04 PM

To: Evans, Barry <Evans.Barry@epa.gov>

Subject: TASK AREA DAS | REGION 7 | CASE 51376 | LAB ACE | SDG GCQ67 & GCQ85 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT |

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

Please advise the laboratory with Case 51376.

ISSUE: "Lab is sending this email with regards to case 51376.

Lab has received water samples for TVOA analysis. Lab has analyzed samples for TVOA analysis for the samples in a continuous analytical sequence where samples are having high concentrations of target analytes and required dilution as well as you can review attached quant reports for the samples. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic

deliverables. Please note that there is no any additional Lab QC failure associated to undiluted and dilution analysis for the samples as per attached.

Please see attached.”

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
information technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>

Sent: Thursday, January 23, 2025 12:34 PM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: TASK AREA DAS | REGION 7 | CASE 51376 | LAB ACE | SDG GCQ67 & GCQ85 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT |

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51376.

Lab has received water samples for TVOA analysis. Lab has analyzed samples for TVOA analysis for the samples in a continuous analytical sequence where samples are having high concentrations of target analytes and required dilution as well as you can review attached quant reports for the samples. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis for final electronic deliverables. Please note that there is no any additional Lab QC failure associated to undiluted and dilution analysis for the samples as per attached.

Please see attached.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

