

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1073 SAS No.: Q1073 SDG No.: Q1073

Instrument ID: MSVOA_L Calibration Date/Time: 01/14/2025 12:55

Lab File ID: VL041889.D Init. Calib. Date(s): 01/14/2025 01/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:40 08:37

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.522	0.466		-10.73	
1,1-Dichloroethene	0.497	0.458		-7.85	
cis-1,2-Dichloroethene	1.282	1.272		-0.78	
1,1,1-Trichloroethane	1.936	1.905		-1.6	
Trichloroethene	0.361	0.354		-1.94	
Tetrachloroethene	0.318	0.292		-8.18	
1-Bromo-4-Fluorobenzene	0.789	0.782		-0.89	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.