

Analysis Method: 9045D
Parameter: pH
Run Number: LB134290
BalanceID: WC SC-7

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.5
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/15/2025	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	01/15/2025	08:06
3	CAL3	1	Water	NA	NA	20.2	10.02	01/15/2025	08:10
4	ICV	1	Water	NA	NA	20.2	7.00	01/15/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	01/15/2025	08:17
6	Q1079-01	1	Solid	20.02	20	20.7	4.16	01/15/2025	08:25
7	Q1079-01DUP	1	Solid	20.03	20	20.8	4.17	01/15/2025	08:26
8	CCV2	1	Water	NA	NA	20.2	12.02	01/15/2025	08:30

WORKLIST(Hardcopy Internal Chain)

NO 134290

WorkList Name : ph s q1079

WorkList ID : 186911

Department : Wet-Chemistry

Date : 01-15-2025 07:48:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1079-01	TAPIAL3-SB04D-R-10-010925-1	Solid	pH	Cool 4 deg C	WEST04	N31	01/09/2025	9045D

Date/Time 01-15-25 07:55
Raw Sample Received by: rd woc
Raw Sample Relinquished by: wjw

Date/Time 01-15-25 10:00
Raw Sample Received by: ef snr
Raw Sample Relinquished by: rd woc