

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1079

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134290

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	01/15/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	01/15/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	01/15/2025

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1079

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134316

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	CCV1						
TOC		mg/L	1040	1000	104	90-110	01/22/2025
Sample ID:	CCV2						
TOC		mg/L	1050	1000	105	90-110	01/22/2025
Sample ID:	ICV1						
TOC		mg/L	934	1000	93	90-110	11/12/2024